


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082020\
 Data File : BN011521.D
 Acq On : 20 Aug 2020 12:30
 Operator : CG/JU
 Sample : DFTPP67
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP67

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081420MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 21 04:54:28 2020



AutoFind: Scans 2431, 2432, 2433; Background Corrected with Scan 2426

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.7	41448	PASS
68	69	0.00	2	1.8	1340	PASS
69	198	0.00	100	37.0	74165	PASS
70	69	0.00	2	0.3	197	PASS
127	198	10	80	46.4	93112	PASS
197	198	0.00	2	0.7	1456	PASS
198	198	100	100	100.0	200640	PASS
199	198	5	9	6.6	13200	PASS
275	198	10	60	27.6	55280	PASS
365	198	1	100	4.3	8577	PASS
441	198	0.01	100	10.9	21907	PASS
442	198	50	100	66.6	133640	PASS
443	442	15	24	21.6	28850	PASS

m/z	Abundance
39.10	454.0
40.00	334.0
41.00	398.0
43.00	631.0
44.00	3720.0
55.10	389.0
57.10	337.0
72.90	827.0
76.90	321.0
83.80	458.0
84.90	429.0
90.90	446.0
147.00	310.0
207.10	1271.0
209.00	391.0
221.00	350.0
281.00	575.0

Instrument :

BNA_N

ClientSampleId :

DFTPP67

m/z	Abundance
37.90	527.0
39.00	3320.0
40.00	518.0
41.10	380.0
43.00	763.0
44.00	3841.0
44.80	569.0
49.10	407.0
50.00	9235.0
51.00	35944.0
52.00	1797.0
55.00	385.0
56.00	952.0
57.00	2853.0
61.00	516.0
62.00	437.0
63.00	1747.0
64.10	520.0
65.00	905.0
68.10	1052.0
68.90	60216.0
69.80	511.0
73.10	1469.0
74.00	8118.0
75.00	10893.0
76.00	4281.0
77.00	68888.0
78.00	4779.0
78.90	5356.0
79.90	4669.0
81.00	5744.0
81.90	1429.0
83.00	1245.0
83.80	519.0
84.90	1521.0
86.00	1634.0
86.90	677.0
90.90	1627.0
91.90	1405.0
93.00	10159.0
94.00	605.0
95.00	317.0
95.80	514.0
96.80	314.0
98.00	8129.0
98.90	5900.0
100.00	538.0
100.90	2634.0
102.90	1153.0
103.90	1554.0
105.00	2116.0
107.00	21952.0
108.00	2919.0
110.00	35936.0
111.00	6122.0
112.10	904.0
115.00	340.0
116.00	1443.0
117.00	25584.0
118.00	2080.0
119.00	398.0
119.90	483.0
121.90	1504.0
122.90	2633.0
123.90	1608.0
125.10	1007.0
126.00	408.0
127.00	69872.0
128.00	5565.0
129.00	29544.0
129.90	2218.0
131.00	466.0
132.10	467.0
134.00	1113.0
134.90	2100.0
135.90	1004.0
137.00	1630.0
139.90	341.0
141.00	4690.0
142.00	1261.0
143.00	1046.0
145.00	314.0
145.90	813.0
147.00	2427.0
148.00	6024.0
148.90	1540.0
151.10	405.0
152.00	418.0
153.00	1216.0
154.10	979.0
154.90	1887.0
156.00	3364.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

157.00	739.0
158.00	830.0
158.90	624.0
159.90	1356.0
160.90	2166.0
161.90	767.0
165.00	1909.0
165.90	1173.0
167.00	8837.0
168.00	6413.0
169.10	834.0
169.80	422.0
170.80	522.0
172.00	765.0
173.00	1010.0
174.00	2161.0
175.00	3275.0
175.90	1029.0
177.00	2100.0
177.90	651.0
179.00	6255.0
179.90	4056.0
181.00	2455.0
182.90	397.0
183.90	812.0
185.00	3050.0
186.00	21304.0
187.00	6615.0
188.00	781.0
189.00	1656.0
190.90	704.0
192.00	2146.0
193.00	2087.0
193.80	749.0
196.00	3285.0
197.10	1037.0
197.90	140928.0
198.90	9253.0
199.90	1168.0
201.50	512.0
202.90	1191.0
204.00	5897.0
205.00	10111.0
206.00	41672.0
207.00	7243.0
208.00	1999.0
208.90	502.0
209.90	719.0
211.00	2085.0
211.80	459.0
215.00	518.0
216.00	1010.0
216.90	12512.0
218.00	1846.0
220.90	4950.0
221.80	1290.0
223.00	2942.0
224.00	21696.0
225.00	5790.0
225.90	1179.0
227.00	10698.0
227.90	2072.0
229.00	2185.0
231.10	895.0
234.00	759.0
234.90	621.0
236.00	856.0
237.00	1198.0
238.90	560.0
240.00	353.0
241.10	546.0
242.00	1531.0
242.90	1244.0
244.00	15829.0
245.00	2966.0
246.00	3461.0
247.00	828.0
248.80	561.0
253.10	628.0
254.90	78688.0
256.00	10843.0
257.00	1328.0
258.00	4317.0
258.80	1154.0
265.00	2201.0
265.90	360.0
270.90	339.0
272.90	2177.0
273.90	7330.0
275.00	35352.0
276.00	5015.0
276.90	4192.0
277.80	734.0
280.90	494.0
282.00	378.0
283.00	467.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

285.10	418.0
293.10	752.0
293.90	416.0
296.00	11083.0
297.00	2310.0
302.90	1497.0
303.90	431.0
313.70	445.0
314.90	1381.0
315.80	641.0
320.80	472.0
322.90	2432.0
324.00	597.0
327.00	683.0
332.10	518.0
333.30	348.0
334.10	1906.0
334.90	614.0
340.90	388.0
345.80	604.0
346.80	327.0
351.90	995.0
352.90	532.0
353.80	967.0
355.00	367.0
365.00	3865.0
365.90	1130.0
370.90	406.0
372.00	1509.0
373.00	431.0
382.90	435.0
402.00	810.0
403.00	821.0
403.90	653.0
421.00	766.0
422.00	536.0
422.90	4672.0
423.80	1833.0
441.00	10713.0
442.00	61016.0
443.00	13696.0
444.00	1185.0

Instrument :

BNA_N

ClientSampleId :

DFTPP67

m/z	Abundance
37.10	439.0
38.00	992.0
39.00	3927.0
39.90	788.0
41.00	559.0
43.00	461.0
44.00	3359.0
45.00	579.0
48.90	560.0
50.00	15743.0
51.00	49888.0
52.00	3084.0
55.00	645.0
55.90	1905.0
57.00	4048.0
57.80	356.0
60.90	978.0
62.00	1603.0
63.00	3061.0
64.10	464.0
65.00	1607.0
66.10	319.0
68.00	1518.0
69.00	91488.0
70.00	591.0
73.00	1952.0
74.00	12031.0
75.00	16656.0
76.00	5503.0
77.00	101784.0
78.00	8282.0
78.90	8195.0
80.00	5012.0
81.00	8443.0
81.90	1790.0
83.00	2112.0
84.00	446.0
85.00	1288.0
85.90	3142.0
86.80	1200.0
87.80	316.0
88.90	418.0
90.90	2549.0
91.90	2496.0
93.00	15716.0
94.00	1029.0
95.00	414.0
96.10	684.0
96.90	370.0
98.00	13366.0
98.90	9968.0
99.90	660.0
101.00	3571.0
101.90	347.0
103.00	1340.0
103.90	3143.0
104.90	3153.0
105.90	919.0
107.00	35616.0
108.00	5332.0
109.10	1368.0
110.00	57824.0
111.00	9686.0
112.00	1456.0
112.90	645.0
115.00	381.0
116.00	2720.0
117.00	42704.0
118.00	3562.0
119.00	320.0
120.00	775.0
122.00	2495.0
123.00	3248.0
123.90	1622.0
124.90	1349.0
126.00	514.0
127.00	111240.0
128.00	8532.0
129.00	49744.0
130.00	4016.0
130.90	1065.0
131.80	469.0
133.00	454.0
133.90	1471.0
135.00	4376.0
136.10	1458.0
136.90	2090.0
138.00	550.0
138.70	337.0
139.90	721.0
140.90	6738.0
142.00	1480.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

143.00	1739.0
143.90	380.0
146.00	1163.0
147.00	3935.0
148.00	10560.0
149.00	1942.0
151.00	779.0
151.90	884.0
153.00	2426.0
153.90	1309.0
155.00	3954.0
156.00	5573.0
157.10	926.0
157.90	1220.0
159.00	737.0
159.90	2435.0
161.00	3178.0
161.80	833.0
163.00	328.0
163.90	387.0
165.00	2573.0
166.10	2224.0
167.00	17048.0
167.90	10122.0
168.90	904.0
170.00	655.0
170.70	544.0
171.90	1629.0
173.00	2238.0
174.00	2937.0
175.00	6472.0
175.90	1013.0
176.20	1059.0
176.90	2487.0
177.90	655.0
179.00	11298.0
180.00	7707.0
181.00	3532.0
181.80	840.0
184.00	789.0
185.00	4648.0
186.00	40112.0
187.00	11135.0
187.90	1460.0
189.00	2599.0
189.90	432.0
191.00	1639.0
191.90	3064.0
193.00	4019.0
194.00	1045.0
195.00	624.0
196.00	7488.0
197.00	1620.0
197.90	232064.0
198.90	16536.0
200.00	1012.0
201.40	931.0
202.90	2343.0
204.00	10672.0
205.00	16776.0
206.00	75880.0
207.00	11755.0
208.00	3442.0
209.00	982.0
210.00	1487.0
211.00	2914.0
211.70	614.0
214.90	1203.0
215.90	2210.0
217.00	22200.0
217.90	2741.0
219.00	429.0
220.20	707.0
221.00	6901.0
223.00	4457.0
224.00	37432.0
225.00	10367.0
226.00	1707.0
227.00	20704.0
228.00	2509.0
229.00	4177.0
230.00	717.0
231.00	1408.0
231.90	355.0
232.90	342.0
234.00	1189.0
234.90	1510.0
235.90	683.0
236.90	1340.0
239.00	765.0
239.90	977.0
240.90	1069.0
242.00	1810.0
242.90	2062.0
244.10	30904.0
244.90	3992.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

245.90	6947.0
247.00	2124.0
249.00	1039.0
250.60	306.0
252.10	372.0
253.00	886.0
254.00	1396.0
255.00	151808.0
256.00	23904.0
257.00	2441.0
257.90	9307.0
259.00	1827.0
259.90	307.0
265.00	4986.0
265.90	554.0
270.00	338.0
271.00	722.0
272.10	539.0
273.00	4631.0
273.90	14135.0
275.00	65344.0
276.00	8301.0
277.00	6938.0
278.00	1656.0
280.90	700.0
281.90	402.0
283.00	1114.0
283.90	538.0
285.00	817.0
292.10	387.0
292.90	1482.0
294.00	457.0
294.90	649.0
296.00	22112.0
297.00	3333.0
301.00	690.0
303.00	2457.0
303.90	511.0
308.00	526.0
314.00	1358.0
314.90	2574.0
316.00	1542.0
320.90	598.0
323.00	5829.0
324.10	968.0
326.80	1274.0
327.90	653.0
332.10	392.0
332.90	1174.0
334.00	3559.0
335.00	1389.0
341.00	797.0
341.90	352.0
346.00	1569.0
346.80	318.0
351.90	1896.0
353.00	1113.0
353.90	2372.0
355.10	539.0
364.90	10874.0
365.90	1422.0
371.20	402.0
372.00	3758.0
373.00	594.0
382.90	751.0
383.80	410.0
389.80	395.0
390.90	305.0
401.90	1678.0
402.90	1731.0
403.80	590.0
420.90	1558.0
422.00	2010.0
423.00	10418.0
424.00	2601.0
425.00	641.0
439.80	403.0
441.00	25280.0
442.00	148928.0
443.00	32048.0
444.00	3116.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
38.00	711.0
39.00	3951.0
40.00	414.0
41.00	311.0
42.00	378.0
43.00	1047.0
44.00	4164.0
44.90	470.0
49.00	533.0
50.00	12748.0
51.00	38512.0
52.00	2034.0
52.90	380.0
55.00	802.0
56.00	1231.0
57.00	3129.0
58.00	345.0
60.00	485.0
61.00	787.0
61.90	1036.0
62.90	2483.0
64.00	404.0
64.90	884.0
68.00	1450.0
68.90	70792.0
72.90	2088.0
73.90	8294.0
75.00	12844.0
76.00	4826.0
77.00	85320.0
78.00	5471.0
79.00	7986.0
80.00	5248.0
80.90	6766.0
81.90	1763.0
82.90	1570.0
83.80	557.0
85.00	1219.0
85.90	2487.0
86.90	1237.0
89.00	353.0
91.00	2423.0
91.90	2266.0
93.00	12422.0
94.00	938.0
95.00	311.0
96.10	681.0
97.10	497.0
98.00	10242.0
98.90	7067.0
100.00	388.0
101.00	3862.0
102.90	1622.0
104.00	2386.0
105.00	2681.0
106.10	1074.0
107.00	29128.0
108.00	4874.0
109.00	1286.0
110.00	47256.0
111.00	7663.0
111.90	1528.0
116.00	1726.0
117.00	35432.0
118.00	2634.0
118.90	433.0
119.80	462.0
122.00	2643.0
123.00	3442.0
124.00	1192.0
125.00	1103.0
126.00	787.0
127.00	98224.0
128.00	7499.0
129.00	41928.0
129.90	3702.0
130.90	785.0
131.90	317.0
132.70	347.0
133.90	1607.0
134.90	3473.0
135.90	1426.0
137.00	2214.0
138.00	600.0
139.90	637.0
141.00	6898.0
142.00	2032.0
142.80	1253.0
143.90	536.0
144.60	334.0
145.90	1249.0
146.90	3324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

147.90	9144.0
148.90	1893.0
149.90	329.0
150.20	353.0
150.80	650.0
151.60	481.0
152.90	1401.0
153.90	1296.0
155.00	3353.0
156.00	4358.0
156.90	1216.0
158.00	1094.0
159.00	1287.0
160.00	1999.0
161.00	3039.0
161.80	764.0
162.90	440.0
164.10	461.0
165.00	2516.0
165.90	1335.0
167.00	14380.0
168.00	7985.0
168.90	863.0
169.90	315.0
170.90	587.0
171.90	1272.0
173.00	1509.0
174.00	2787.0
175.00	4985.0
175.90	1367.0
177.00	2382.0
177.90	892.0
179.00	9915.0
180.00	6686.0
181.00	2792.0
181.90	721.0
183.80	788.0
185.00	6648.0
186.00	35320.0
187.00	9827.0
187.90	982.0
188.90	2424.0
190.10	512.0
190.90	1500.0
191.90	3177.0
193.00	4604.0
194.20	628.0
195.00	481.0
195.90	6832.0
197.10	1711.0
198.00	228928.0
198.90	13812.0
199.90	1776.0
200.70	662.0
201.30	922.0
202.00	479.0
203.00	2108.0
204.00	10608.0
205.00	17120.0
206.00	67008.0
207.00	10975.0
208.00	2911.0
208.90	914.0
210.10	1135.0
211.00	3438.0
212.10	410.0
215.00	1113.0
216.10	2031.0
217.00	19096.0
217.90	2779.0
219.00	479.0
220.10	532.0
220.90	8810.0
221.70	2837.0
223.00	4209.0
224.00	37768.0
225.00	10763.0
226.10	1195.0
227.00	18568.0
227.90	2679.0
229.00	3170.0
229.90	527.0
231.00	1261.0
233.90	978.0
235.00	916.0
236.00	1222.0
236.90	1830.0
238.90	608.0
240.10	702.0
241.00	764.0
242.00	1779.0
243.10	2370.0
244.00	29008.0
245.00	4425.0
246.00	6709.0
247.00	1254.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

247.90	590.0
248.80	1173.0
251.10	358.0
251.80	531.0
253.00	790.0
254.10	1432.0
255.00	171776.0
256.00	24904.0
257.00	2291.0
258.00	11132.0
259.00	1851.0
259.90	419.0
265.00	3734.0
266.00	777.0
270.00	313.0
271.00	528.0
272.20	507.0
272.90	4544.0
274.00	11348.0
275.00	65144.0
276.00	8980.0
277.00	6488.0
277.90	1299.0
279.00	326.0
280.90	832.0
281.90	519.0
283.00	973.0
283.80	537.0
284.90	845.0
292.00	351.0
293.00	1322.0
294.10	475.0
296.00	25280.0
297.00	2924.0
300.90	437.0
301.90	617.0
302.90	1878.0
304.00	811.0
309.00	326.0
310.00	357.0
313.00	463.0
314.00	1084.0
315.00	2834.0
316.00	1557.0
316.90	390.0
321.00	852.0
323.00	5895.0
324.10	1300.0
326.60	1133.0
327.90	1020.0
332.10	509.0
333.10	791.0
334.00	4969.0
334.90	1198.0
341.10	1075.0
345.90	1232.0
352.00	2178.0
352.80	1241.0
354.00	2028.0
354.90	879.0
364.90	10993.0
366.00	1708.0
371.00	322.0
372.00	3998.0
372.90	871.0
382.90	1139.0
390.00	731.0
390.90	624.0
401.90	1575.0
402.90	2319.0
404.00	1127.0
421.00	2091.0
422.00	1945.0
423.00	13002.0
424.00	3499.0
441.00	29728.0
442.00	190976.0
443.00	40808.0
443.90	4063.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67