



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082021\
 Data File : BN015982.D
 Acq On : 20 Aug 2021 11:25
 Operator : CG/JU
 Sample : DFTPP203
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP203

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 20 04:24:09 2021



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	433464	PASS
68	69	0.00	2	2.0	9779	PASS
69	198	0.00	100	43.2	489856	PASS
70	69	0.00	2	0.5	2522	PASS
127	198	10	80	46.2	524608	PASS
197	198	0.00	2	0.3	2971	PASS
198	198	100	100	100.0	1134656	PASS
199	198	5	9	6.9	78797	PASS
275	198	10	60	29.0	328811	PASS
365	198	1	100	3.6	40528	PASS
441	198	0.01	100	16.1	183016	PASS
442	198	50	100	98.4	1116501	PASS
443	442	15	24	20.0	222805	PASS

m/z	Abundance
40.00	1265.0
41.10	489.0
42.90	611.0
43.90	2243.0
48.90	845.0
51.00	520.0
57.10	350.0
68.80	309.0
73.10	2300.0
76.00	338.0
82.80	349.0
83.90	961.0
86.00	418.0
95.80	356.0
132.80	343.0
147.00	655.0
167.00	566.0
176.80	341.0
191.00	504.0
193.10	339.0
207.00	2940.0
208.00	659.0
209.10	872.0
221.10	378.0
281.10	1252.0
282.90	369.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
36.00	562.0
37.00	2153.0
38.10	5117.0
39.10	29344.0
40.00	2407.0
41.10	1230.0
43.10	509.0
44.00	2359.0
45.00	690.0
47.00	389.0
49.00	3893.0
50.00	109008.0
51.00	391168.0
52.00	19800.0
53.10	1107.0
55.00	2708.0
56.00	13217.0
57.00	27920.0
57.90	1020.0
58.80	380.0
61.00	6218.0
62.00	6343.0
63.00	17936.0
64.00	2196.0
65.10	6898.0
66.10	709.0
66.90	778.0
68.10	10344.0
69.00	444224.0
70.00	2579.0
72.10	602.0
73.00	4724.0
74.00	45520.0
75.00	69456.0
76.10	21880.0
77.00	460480.0
78.10	30480.0
79.00	33952.0
80.00	26272.0
81.00	36168.0
82.00	9027.0
82.90	7302.0
84.00	1617.0
84.90	6614.0
86.00	10624.0
87.00	4169.0
88.00	1689.0
89.00	1019.0
91.00	8047.0
91.90	9116.0
93.00	55912.0
93.90	3225.0
95.00	800.0
96.00	2550.0
97.10	987.0
98.00	42688.0
99.00	35192.0
100.00	2484.0
101.00	19872.0
101.90	874.0
103.00	6887.0
104.00	11279.0
105.00	12219.0
106.10	3633.0
107.00	137024.0
108.00	21560.0
109.00	4121.0
110.00	252160.0
111.00	38392.0
112.00	4686.0
113.00	1657.0
115.00	1085.0
116.00	9007.0
117.00	119720.0
118.00	9287.0
118.90	1656.0
120.00	1547.0
120.90	875.0
122.00	9222.0
123.00	12852.0
124.00	6588.0
125.00	6258.0
127.00	459136.0
128.00	33656.0
129.00	182272.0
130.00	14732.0
130.90	2769.0
132.00	2227.0
133.00	1000.0
134.00	5024.0
135.00	15578.0
136.00	6487.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

137.00	6983.0
138.00	2074.0
138.90	873.0
139.90	2573.0
141.00	23392.0
142.10	7547.0
142.90	5281.0
143.90	1802.0
144.90	1280.0
145.90	4675.0
147.00	13319.0
148.00	27584.0
149.00	6926.0
150.00	1653.0
151.10	3270.0
151.80	1555.0
153.10	9033.0
154.00	6124.0
155.00	13220.0
156.00	18520.0
157.10	4097.0
158.00	3781.0
159.00	2972.0
160.00	7901.0
161.00	11939.0
161.80	3315.0
163.00	1032.0
163.80	1023.0
164.90	8584.0
166.00	6624.0
167.00	44160.0
168.00	21232.0
169.00	3982.0
170.10	1299.0
170.90	2941.0
171.90	4758.0
173.00	6160.0
174.00	9423.0
175.10	19192.0
176.10	4809.0
177.00	9411.0
178.00	3048.0
179.00	33920.0
180.00	20936.0
181.10	11363.0
182.00	1575.0
182.70	867.0
184.00	2551.0
185.00	18616.0
186.00	129584.0
187.00	37584.0
188.20	4014.0
189.00	8625.0
190.00	1949.0
191.00	4115.0
192.10	11118.0
193.00	11191.0
194.00	2681.0
195.00	1605.0
196.00	24072.0
197.10	8912.0
197.90	902336.0
199.00	62680.0
200.10	4317.0
200.80	948.0
201.50	3198.0
203.00	6561.0
204.00	31616.0
205.00	57384.0
206.00	229952.0
207.00	36976.0
208.00	8987.0
209.00	3050.0
210.10	3614.0
211.00	8883.0
213.00	803.0
215.00	2360.0
216.00	4820.0
217.00	65840.0
218.00	7839.0
218.90	1183.0
219.80	1112.0
221.00	36248.0
221.80	10119.0
223.00	13940.0
224.00	138752.0
225.10	33376.0
226.00	4773.0
227.00	56448.0
228.00	7612.0
228.90	9938.0
229.90	1092.0
231.00	4249.0
232.10	821.0
233.00	842.0
233.90	3205.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

235.00	3995.0
236.10	1932.0
236.90	5126.0
237.90	1063.0
239.10	2233.0
240.10	1864.0
240.90	3134.0
242.00	6215.0
243.10	8270.0
244.00	98736.0
245.10	13831.0
246.00	21968.0
247.00	4875.0
247.90	1000.0
249.10	3585.0
249.90	958.0
251.00	937.0
252.00	1263.0
253.00	2048.0
255.00	545600.0
256.10	81048.0
257.00	5714.0
258.00	28904.0
258.90	4956.0
260.00	705.0
261.00	1076.0
263.00	606.0
263.90	803.0
265.00	12137.0
266.00	1791.0
267.10	710.0
269.80	614.0
271.20	954.0
272.00	1270.0
273.00	16193.0
274.00	41352.0
275.00	242432.0
276.00	33400.0
277.00	21416.0
278.00	4247.0
279.00	1036.0
281.10	996.0
282.00	728.0
282.90	1845.0
283.90	1756.0
285.00	4091.0
286.00	623.0
288.00	422.0
289.00	988.0
290.10	985.0
291.00	554.0
292.10	770.0
293.00	4435.0
294.00	1048.0
296.00	60752.0
296.90	9062.0
298.10	542.0
299.10	419.0
300.80	398.0
301.10	378.0
301.80	865.0
302.30	658.0
303.00	8543.0
304.10	1993.0
307.90	1065.0
308.90	736.0
309.90	979.0
313.10	751.0
314.00	3397.0
315.00	6737.0
316.00	3829.0
317.10	480.0
319.90	324.0
321.10	1445.0
322.30	1417.0
323.00	21808.0
324.00	2968.0
325.00	621.0
325.90	495.0
327.00	4260.0
327.90	1643.0
329.00	556.0
332.00	1945.0
333.00	1905.0
334.00	14572.0
335.00	3616.0
336.00	407.0
338.90	486.0
341.00	2625.0
342.00	1034.0
345.90	4699.0
347.00	568.0
352.00	6412.0
353.10	4587.0
354.00	7017.0
355.10	1233.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

359.10	357.0
365.00	24640.0
366.00	4659.0
370.00	521.0
370.90	1095.0
372.10	11850.0
373.00	2390.0
373.90	486.0
382.90	3592.0
383.90	1145.0
384.80	616.0
390.00	1637.0
390.80	596.0
391.30	794.0
392.10	644.0
400.90	798.0
402.00	4375.0
403.00	6578.0
404.00	2543.0
405.00	497.0
415.00	515.0
421.00	5233.0
422.00	5333.0
422.90	40720.0
424.00	11506.0
425.00	2009.0
441.00	103096.0
442.00	707072.0
443.00	145088.0
444.00	14630.0
444.90	661.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
36.00	392.0
36.90	2266.0
38.10	6501.0
39.10	39240.0
40.00	2763.0
41.00	1449.0
43.00	341.0
44.00	2255.0
45.10	1149.0
46.80	348.0
48.10	444.0
49.00	4867.0
50.00	136896.0
51.00	508608.0
52.00	25472.0
53.00	1149.0
55.10	3351.0
56.00	17112.0
57.00	35248.0
58.00	1251.0
58.90	604.0
59.90	331.0
61.00	6563.0
62.00	8271.0
63.00	21816.0
64.00	3164.0
65.00	8888.0
65.90	835.0
67.10	730.0
68.10	10803.0
69.00	567168.0
70.00	2849.0
71.00	531.0
71.90	359.0
73.00	5371.0
74.00	58376.0
75.00	88936.0
76.00	28832.0
77.00	577216.0
78.00	41736.0
79.00	44456.0
80.00	33320.0
81.00	44408.0
82.00	10683.0
82.90	10017.0
84.10	1735.0
85.00	9005.0
86.00	15409.0
87.00	6118.0
88.00	2246.0
89.00	983.0
91.00	10188.0
92.00	12305.0
93.00	73728.0
94.00	4967.0
94.90	1158.0
96.10	3616.0
97.10	1799.0
98.00	58752.0
99.00	45160.0
100.00	4066.0
101.00	23496.0
102.10	1041.0
103.00	9223.0
104.00	17208.0
105.00	15987.0
106.00	5375.0
107.00	190336.0
108.00	29088.0
109.10	6178.0
110.00	331072.0
111.00	54912.0
112.00	7599.0
113.00	2120.0
114.10	456.0
115.00	1212.0
116.00	11855.0
117.00	156352.0
118.00	11140.0
119.10	1205.0
120.10	2050.0
120.80	1110.0
122.00	13798.0
123.00	19024.0
124.00	8722.0
125.00	6779.0
127.00	601152.0
128.00	45784.0
129.00	253312.0
130.00	22040.0
131.00	4780.0
132.00	1666.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

133.00	1185.0
133.90	8197.0
135.00	19664.0
136.00	8549.0
137.00	9835.0
138.10	2473.0
139.00	1278.0
140.00	3669.0
141.00	33848.0
142.00	10415.0
143.00	6419.0
144.00	1831.0
145.00	1806.0
146.00	5452.0
147.00	14598.0
148.00	37520.0
149.00	7785.0
150.00	2018.0
151.10	3600.0
151.90	2816.0
153.00	10886.0
154.00	8847.0
155.00	18688.0
156.10	26600.0
157.00	4483.0
158.00	5710.0
158.90	4598.0
160.00	9818.0
161.00	15018.0
162.10	3822.0
163.00	1271.0
164.10	1526.0
165.00	12450.0
166.00	9496.0
167.00	62200.0
168.00	30384.0
169.10	5072.0
170.00	2241.0
171.00	2754.0
171.90	6578.0
173.00	7074.0
174.00	12621.0
175.00	25320.0
176.00	7703.0
176.90	13306.0
178.00	4506.0
179.00	48448.0
180.00	33056.0
181.00	14908.0
182.00	2257.0
183.10	949.0
184.00	3483.0
185.00	24112.0
186.00	190592.0
187.00	56400.0
188.00	4376.0
189.00	12281.0
190.00	2332.0
191.10	5272.0
192.00	15675.0
193.00	16456.0
194.00	4017.0
196.00	38736.0
198.00	1329152.0
199.00	92904.0
200.00	7227.0
201.40	6014.0
202.20	1682.0
203.00	10387.0
204.00	51304.0
205.00	87664.0
206.00	366784.0
207.10	51920.0
208.00	12866.0
209.00	4197.0
210.20	5874.0
211.00	13102.0
213.00	818.0
214.10	431.0
215.00	4699.0
216.10	8475.0
217.00	95568.0
218.00	13017.0
218.90	1372.0
221.00	52960.0
221.80	11559.0
223.00	21448.0
224.00	201600.0
225.10	55312.0
226.10	6480.0
227.00	85936.0
228.00	12962.0
229.00	16944.0
230.00	2315.0
231.00	6545.0
231.90	891.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

233.10	1805.0
234.00	5234.0
235.00	5531.0
236.00	4199.0
237.00	6918.0
237.90	1162.0
238.10	1166.0
239.10	2984.0
240.10	3271.0
240.90	4657.0
242.00	8087.0
243.00	10500.0
244.00	151488.0
245.10	20080.0
246.00	30456.0
247.00	5494.0
247.90	1409.0
249.00	5925.0
250.00	1302.0
251.00	1557.0
252.10	1257.0
253.00	2678.0
255.00	795008.0
256.00	124928.0
257.00	10204.0
258.00	44608.0
259.00	7842.0
260.10	1602.0
260.80	1915.0
263.00	599.0
263.80	1232.0
265.00	18744.0
266.00	2715.0
267.10	579.0
268.00	489.0
270.10	1028.0
270.90	1735.0
272.10	2271.0
273.00	23480.0
274.00	62592.0
275.00	380416.0
276.00	48184.0
277.00	31472.0
278.00	5078.0
278.90	1111.0
281.10	1065.0
281.90	1221.0
283.00	4012.0
284.00	2746.0
285.10	6700.0
286.00	1235.0
287.90	490.0
288.90	1176.0
290.10	828.0
290.90	822.0
292.10	1771.0
293.00	7944.0
294.00	1638.0
296.00	99960.0
297.00	16010.0
298.10	1011.0
301.10	1340.0
302.00	2059.0
303.00	11180.0
303.90	3663.0
305.00	560.0
308.10	1820.0
308.90	1134.0
309.90	990.0
311.10	577.0
312.00	431.0
312.90	1103.0
314.00	5016.0
315.00	11199.0
316.00	5500.0
317.00	1678.0
319.80	464.0
320.90	2483.0
321.90	2087.0
323.00	32296.0
324.00	6036.0
325.00	898.0
325.90	993.0
327.00	6881.0
328.00	2702.0
329.00	648.0
331.90	2515.0
333.10	2602.0
334.00	19584.0
335.10	4673.0
335.90	959.0
339.80	626.0
340.30	465.0
341.00	3723.0
342.00	1073.0
345.90	6705.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

347.00	1230.0
350.80	654.0
352.10	12007.0
353.10	6600.0
354.00	9520.0
354.90	1893.0
358.80	338.0
365.00	48696.0
366.00	6004.0
367.00	803.0
370.10	1549.0
371.00	2566.0
372.00	19360.0
373.00	3956.0
373.90	608.0
376.90	759.0
382.90	5091.0
383.90	1958.0
385.00	425.0
389.90	2511.0
391.00	1708.0
392.20	1488.0
401.10	917.0
401.90	7862.0
402.90	10862.0
403.90	4161.0
404.90	731.0
415.00	578.0
421.00	9908.0
421.90	8578.0
423.00	71840.0
424.00	18544.0
425.10	3022.0
441.00	209280.0
442.00	1240064.0
443.00	245504.0
444.00	24104.0
445.00	1189.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
36.00	329.0
37.00	1272.0
38.00	6051.0
39.10	27368.0
40.00	2078.0
40.90	1105.0
43.00	441.0
44.00	2091.0
45.10	1003.0
46.90	381.0
48.00	346.0
49.10	4186.0
50.00	108712.0
51.00	402176.0
52.00	20136.0
53.00	879.0
55.00	1815.0
56.00	11990.0
57.00	28840.0
57.90	1371.0
59.00	641.0
60.10	370.0
61.00	5932.0
62.00	7193.0
63.00	18632.0
64.10	2535.0
65.00	7649.0
65.90	401.0
67.10	746.0
68.10	8191.0
69.00	458176.0
70.00	2139.0
71.10	602.0
73.00	5251.0
74.00	48384.0
75.00	75512.0
76.00	24656.0
77.00	468480.0
78.00	32832.0
79.00	38616.0
80.00	26576.0
81.00	37936.0
82.00	8880.0
83.10	7640.0
84.00	1973.0
85.00	7509.0
86.00	10731.0
87.00	6072.0
88.00	1837.0
89.00	688.0
90.90	8870.0
92.00	9317.0
93.00	62624.0
94.00	4327.0
95.00	1089.0
96.10	2470.0
97.10	1705.0
98.00	51064.0
99.00	35696.0
100.00	2895.0
101.00	19912.0
102.00	891.0
103.00	8293.0
104.00	13212.0
105.00	12133.0
106.00	4396.0
107.00	155712.0
108.00	23400.0
109.10	4974.0
110.00	283520.0
111.00	44008.0
112.00	5686.0
113.10	1380.0
113.90	376.0
115.00	936.0
116.00	8754.0
117.00	135296.0
118.00	9666.0
119.10	1506.0
120.00	2066.0
121.00	864.0
122.00	12549.0
123.00	16944.0
124.00	7847.0
125.00	7249.0
127.00	513536.0
128.00	39856.0
129.00	219328.0
130.00	18984.0
131.00	3406.0
132.10	2003.0
132.90	1110.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

134.00	5594.0
135.00	17800.0
136.00	7821.0
137.00	8430.0
138.10	1864.0
138.90	1036.0
140.00	2851.0
141.00	27216.0
142.00	11234.0
142.90	5350.0
144.10	1614.0
145.00	1183.0
146.00	5708.0
147.00	15873.0
148.00	35368.0
149.00	7120.0
150.00	2396.0
151.10	3039.0
151.90	2770.0
153.00	9821.0
154.00	6557.0
155.00	16488.0
156.00	23152.0
157.00	5232.0
157.90	5928.0
159.10	4428.0
160.00	9795.0
161.00	13236.0
162.00	3509.0
163.00	1314.0
164.00	1434.0
165.00	11209.0
166.00	9394.0
167.00	53104.0
168.00	25392.0
169.10	4485.0
169.90	2141.0
171.00	2159.0
172.00	6204.0
173.00	7593.0
174.10	12621.0
175.10	21328.0
176.00	5537.0
177.00	14169.0
178.10	3472.0
179.00	44336.0
180.00	29000.0
181.00	14207.0
181.90	1757.0
182.80	1226.0
183.90	3362.0
185.00	22968.0
186.00	169088.0
187.00	47528.0
188.00	3799.0
189.10	11151.0
190.00	1874.0
191.10	5422.0
192.00	14626.0
193.00	15715.0
194.10	3751.0
195.20	1335.0
196.00	33304.0
198.00	1172480.0
199.00	80808.0
200.00	5826.0
201.50	5778.0
202.20	811.0
203.00	9153.0
204.00	45472.0
205.00	72456.0
206.10	323328.0
207.10	45040.0
208.00	11858.0
209.00	4087.0
210.20	5844.0
211.10	11034.0
211.90	2084.0
213.10	802.0
214.00	469.0
215.00	3781.0
216.00	8761.0
217.00	88248.0
218.00	10296.0
219.00	1320.0
221.00	48888.0
223.00	20496.0
224.00	194112.0
225.00	44632.0
226.00	5089.0
227.00	79408.0
228.00	11206.0
229.00	16752.0
229.90	2139.0
231.00	6562.0
232.00	1100.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

233.00	1598.0
233.90	4645.0
235.00	5310.0
236.00	4328.0
236.90	5607.0
238.10	963.0
239.00	2755.0
240.00	2328.0
241.00	4008.0
242.00	11128.0
243.10	9638.0
244.00	142016.0
245.10	16720.0
246.00	30064.0
247.00	5788.0
247.90	1290.0
249.00	6399.0
249.90	1223.0
251.00	1999.0
252.10	2003.0
252.90	4063.0
254.20	10082.0
255.00	771840.0
256.00	111088.0
257.00	9633.0
258.00	46800.0
259.10	7393.0
260.00	1396.0
260.90	1384.0
263.00	637.0
264.00	919.0
265.00	20040.0
266.00	2668.0
267.00	374.0
268.00	336.0
269.90	716.0
270.90	1462.0
272.00	1830.0
273.00	24592.0
274.00	64248.0
275.00	363584.0
276.00	43304.0
277.00	31360.0
278.10	5618.0
279.00	670.0
281.00	1304.0
282.10	979.0
282.90	3093.0
283.90	3167.0
285.00	5115.0
286.10	648.0
288.10	531.0
289.00	1155.0
290.10	851.0
291.00	1233.0
292.00	1380.0
293.00	7344.0
294.00	1448.0
295.10	2612.0
296.00	101104.0
297.00	15490.0
298.10	787.0
299.00	470.0
300.90	1064.0
302.00	1558.0
303.00	10844.0
304.00	3608.0
304.90	500.0
308.00	1617.0
308.80	1075.0
309.90	1503.0
312.80	798.0
314.00	6081.0
315.00	13069.0
316.00	7425.0
317.00	1000.0
321.10	3267.0
323.00	32640.0
324.00	5668.0
324.90	678.0
326.00	759.0
327.00	5961.0
328.00	3224.0
329.00	581.0
332.10	2725.0
332.90	2879.0
334.00	23264.0
335.00	4637.0
336.00	568.0
338.70	436.0
339.10	383.0
339.80	349.0
341.00	4009.0
341.90	1232.0
345.90	6958.0
346.90	1138.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

349.90	548.0
350.90	579.0
352.00	10450.0
352.90	6600.0
354.00	11560.0
355.10	2425.0
356.00	345.0
357.10	380.0
359.00	1014.0
359.80	330.0
365.00	48248.0
366.00	6692.0
367.00	475.0
369.90	1161.0
371.10	3238.0
372.00	19224.0
373.00	5075.0
374.00	702.0
377.10	313.0
382.90	5583.0
383.90	1890.0
384.80	675.0
390.10	2194.0
391.00	2114.0
392.00	1710.0
400.90	1105.0
401.90	7602.0
402.90	12917.0
403.90	4523.0
405.10	983.0
409.90	430.0
414.90	709.0
421.00	11051.0
422.10	11350.0
423.00	82528.0
424.00	17848.0
425.00	1675.0
441.00	236672.0
442.00	1402368.0
443.00	277824.0
444.00	26224.0
445.00	1423.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203