

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082024\
 Data File : BN033514.D
 Acq On : 20 Aug 2024 20:04
 Operator : MA/JU
 Sample : P3646-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 916-J-WPO-0.25-081524

Quant Time: Aug 20 23:22:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 23:32:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.552	152	8482	0.400	ng	0.00
7) Naphthalene-d8	10.314	136	22740	0.400	ng	# 0.00
13) Acenaphthene-d10	14.189	164	11052	0.400	ng	0.00
19) Phenanthrene-d10	16.929	188	22302	0.400	ng	#-0.01
29) Chrysene-d12	21.148	240	14089	0.400	ng	# 0.00
35) Perylene-d12	23.315	264	13411	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.183	112	3340	0.124	ng	0.00
5) Phenol-d6	6.736	99	2750	0.086	ng	0.00
8) Nitrobenzene-d5	8.681	82	4680	0.248	ng	-0.01
11) 2-Methylnaphthalene-d10	11.907	152	8410	0.259	ng	0.00
14) 2,4,6-Tribromophenol	15.688	330	1436	0.242	ng	0.00
15) 2-Fluorobiphenyl	12.810	172	13608	0.301	ng	0.00
27) Fluoranthene-d10	18.975	212	15848	0.296	ng	0.00
31) Terphenyl-d14	19.593	244	10852	0.339	ng	0.00
Target Compounds						
9) Naphthalene	10.357	128	1717	0.028	ng	# 86
17) Acenaphthene	14.242	154	3641	0.107	ng	95
18) Fluorene	15.236	166	3839	0.089	ng	# 63
24) Pentachlorophenol	16.594	266	287	0.044	ng	96
25) Phenanthrene	16.979	178	1465	0.024	ng	# 92
28) Fluoranthene	19.007	202	3177	0.046	ng	99
30) Pyrene	19.370	202	1792	0.028	ng	# 95
33) Chrysene	21.175	228	1161	0.023	ng	# 70
34) Bis(2-ethylhexyl)phtha...	21.095	149	2379	0.074	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082024\
Data File : BN033514.D
Acq On : 20 Aug 2024 20:04
Operator : MA/JU
Sample : P3646-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
916-J-WPO-0.25-081524

Quant Time: Aug 20 23:22:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 23:32:18 2024
Response via : Initial Calibration

