

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082024\  
 Data File : BN033519.D  
 Acq On : 20 Aug 2024 23:06  
 Operator : MA/JU  
 Sample : P3666-02  
 Misc :  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SP-201-20240816

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 08/21/2024  
 Supervised By :mohammad ahmed 08/22/2024

Quant Time: Aug 20 23:33:24 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN081924.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 19 23:32:18 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.552  | 152  | 6956     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.314 | 136  | 17989    | 0.400 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.189 | 164  | 8326     | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 16.929 | 188  | 15938    | 0.400 | ng    | #-0.01   |
| 29) Chrysene-d12              | 21.139 | 240  | 10259    | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.312 | 264  | 10334    | 0.400 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.183  | 112  | 2300     | 0.104 | ng    | 0.00     |
| 5) Phenol-d6                  | 6.736  | 99   | 1625     | 0.062 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 8.681  | 82   | 4846     | 0.325 | ng    | -0.01    |
| 11) 2-Methylnaphthalene-d10   | 11.907 | 152  | 7657     | 0.298 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.688 | 330  | 978      | 0.219 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 12.809 | 172  | 11526    | 0.339 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 18.975 | 212  | 13818    | 0.361 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.593 | 244  | 10940    | 0.469 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
| 34) Bis(2-ethylhexyl)phtha... | 21.086 | 149  | 1269m    | 0.054 | ng    | Qvalue   |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082024\  
Data File : BN033519.D  
Acq On : 20 Aug 2024 23:06  
Operator : MA/JU  
Sample : P3666-02  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SP-201-20240816

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 08/21/2024  
Supervised By :mohammad ahmed 08/22/2024

Quant Time: Aug 20 23:33:24 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN081924.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Aug 19 23:32:18 2024  
Response via : Initial Calibration

