

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
 Data File : BN037637.D
 Acq On : 20 Aug 2025 19:22
 Operator : RC/JU
 Sample : Q2885-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW5-SP100-20250814

Quant Time: Aug 21 04:57:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	2105	0.400	ng	0.00
7) Naphthalene-d8	10.487	136	4906	0.400	ng	#-0.01
13) Acenaphthene-d10	14.345	164	2332	0.400	ng	0.00
19) Phenanthrene-d10	17.086	188	5055	0.400	ng	# 0.00
29) Chrysene-d12	21.268	240	4261	0.400	ng	# 0.00
35) Perylene-d12	23.499	264	3736	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.305	112	631	0.132	ng	0.00
5) Phenol-d6	6.879	99	426	0.074	ng	0.00
8) Nitrobenzene-d5	8.854	82	1237	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	12.085	152	1861	0.279	ng	0.00
14) 2,4,6-Tribromophenol	15.833	330	371	0.364	ng	0.00
15) 2-Fluorobiphenyl	12.968	172	4388	0.325	ng	-0.01
27) Fluoranthene-d10	19.113	212	4911	0.369	ng	0.00
31) Terphenyl-d14	19.717	244	3906	0.446	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.232	88	11309	5.615	ng	98
6) bis(2-Chloroethyl)ether	7.132	93	200	0.039	ng	# 78
24) Pentachlorophenol	16.739	266	55	0.035	ng	96
34) Bis(2-ethylhexyl)phtha...	21.196	149	1455	0.248	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
Data File : BN037637.D
Acq On : 20 Aug 2025 19:22
Operator : RC/JU
Sample : Q2885-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW5-SP100-20250814

Quant Time: Aug 21 04:57:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 13 05:00:58 2025
Response via : Initial Calibration

