




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
Data File : BN037632.D
Acq On : 20 Aug 2025 16:18
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 21 05:59:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 06 01:49:15 2025
Response via : Initial Calibration



TIC: BN037632.D\data.ms

(70) Pentachlorophenol (C)




16.734min (-0.012) 25391.22 ng

response 190297

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.37
264.00	61.60	58.51
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
Data File : BN037632.D
Acq On : 20 Aug 2025 16:18
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 21 05:59:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 06 01:49:15 2025
Response via : Initial Calibration



TIC: BN037632.D\data.ms

(77) Benzidine

19.334min (-0.001) 0.00 ng m

response 940808

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.76
183.00	13.20	12.63
0.00	0.00	0.00