

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
 Data File : BN037641.D
 Acq On : 20 Aug 2025 21:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 21 04:58:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	122	0.00
2	1,4-Dioxane	0.400	0.366	8.5	108	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	127	0.00
4 S	2-Fluorophenol	0.400	0.493	-23.2	154	0.00
5 S	Phenol-d6	0.400	0.782	-95.5#	250	0.00
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	134	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	-0.01
8 S	Nitrobenzene-d5	0.400	0.539	-34.8#	162	0.00
9	Naphthalene	0.400	0.430	-7.5	123	-0.01
10	Hexachlorobutadiene	0.400	0.460	-15.0	127	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.340	15.0	102	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.407	-1.7	118	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	105	0.00
16	Acenaphthylene	0.400	0.387	3.3	105	-0.01
17	Acenaphthene	0.400	0.384	4.0	104	-0.01
18	Fluorene	0.400	0.392	2.0	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.390	2.5	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.380	5.0	110	-0.01
22	Hexachlorobenzene	0.400	0.411	-2.7	111	-0.01
23	Atrazine	0.400	0.427	-6.7	126	0.00
24	Pentachlorophenol	0.400	0.367	8.3	109	0.00
25	Phenanthrene	0.400	0.383	4.3	107	0.00
26	Anthracene	0.400	0.372	7.0	108	-0.01
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	109	0.00
28	Fluoranthene	0.400	0.373	6.8	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	108	0.00
30	Pyrene	0.400	0.396	1.0	109	0.00
31 S	Terphenyl-d14	0.400	0.393	1.8	110	0.00
32	Benzo(a)anthracene	0.400	0.377	5.8	108	0.00
33	Chrysene	0.400	0.390	2.5	110	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.415	-3.7	119	0.00
35 I	Perylene-d12	0.400	0.400	0.0	99	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	99	-0.02
37	Benzo(b)fluoranthene	0.400	0.415	-3.7	116	-0.01
38	Benzo(k)fluoranthene	0.400	0.380	5.0	97	0.00
39 C	Benzo(a)pyrene	0.400	0.379	5.3	103	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.375	6.3	99	-0.02
41	Benzo(g,h,i)perylene	0.400	0.403	-0.8	109	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4