

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
 Data File : BN037643.D
 Acq On : 21 Aug 2025 09:51
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 22 08:22:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.383	0.382	0.3	106	0.00
3	n-Nitrosodimethylamine	0.489	0.495	-1.2	111	0.00
4 S	2-Fluorophenol	0.907	0.826	8.9	102	0.00
5 S	Phenol-d6	1.091	1.009	7.5	106	0.00
6	bis(2-Chloroethyl)ether	0.983	0.965	1.8	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.01
8 S	Nitrobenzene-d5	0.282	0.277	1.8	112	-0.01
9	Naphthalene	1.065	1.059	0.6	108	-0.01
10	Hexachlorobutadiene	0.260	0.264	-1.5	107	-0.01
11 SURR	2-Methylnaphthalene-d10	0.544	0.487	10.5	102	-0.01
12	2-Methylnaphthalene	0.668	0.633	5.2	106	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.175	0.178	-1.7	117	0.00
15 S	2-Fluorobiphenyl	2.313	2.301	0.5	103	-0.01
16	Acenaphthylene	1.793	1.783	0.6	107	-0.01
17	Acenaphthene	1.220	1.201	1.6	106	-0.01
18	Fluorene	1.596	1.609	-0.8	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.039	26.4#	93	0.00
21	4-Bromophenyl-phenylether	0.251	0.238	5.2	111	-0.01
22	Hexachlorobenzene	0.356	0.369	-3.7	113	-0.01
23	Atrazine	0.142	0.145	-2.1	121	0.00
24	Pentachlorophenol	0.125	0.252	-101.6#	242#	0.00
25	Phenanthrene	1.216	1.172	3.6	109	0.00
26	Anthracene	1.077	1.003	6.9	110	-0.01
27 SURR	Fluoranthene-d10	1.052	0.955	9.2	109	0.00
28	Fluoranthene	1.397	1.295	7.3	109	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
30	Pyrene	1.495	1.463	2.1	109	0.00
31 S	Terphenyl-d14	0.823	0.790	4.0	108	0.00
32	Benzo(a)anthracene	1.336	1.279	4.3	111	0.00
33	Chrysene	1.489	1.471	1.2	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.551	0.593	-7.6	124	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	1.686	1.575	6.6	94	-0.01
37	Benzo(b)fluoranthene	1.516	1.560	-2.9	112	-0.01
38	Benzo(k)fluoranthene	1.710	1.553	9.2	90	0.00
39 C	Benzo(a)pyrene	1.257	1.188	5.5	99	0.00
40	Dibenzo(a,h)anthracene	1.308	1.165	10.9	92	-0.01
41	Benzo(g,h,i)perylene	1.378	1.414	-2.6	108	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4