

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082025\
 Data File : BN037646.D
 Acq On : 21 Aug 2025 13:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 22 08:23:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	-0.01
2	1,4-Dioxane	0.400	0.408	-2.0	113	0.00
3	n-Nitrosodimethylamine	0.400	0.399	0.3	114	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	108	-0.01
5 S	Phenol-d6	0.400	0.358	10.5	107	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	113	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.400	0.400	0.0	117	-0.01
9	Naphthalene	0.400	0.393	1.8	110	-0.02
10	Hexachlorobutadiene	0.400	0.411	-2.7	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.361	9.8	105	-0.01
12	2-Methylnaphthalene	0.400	0.373	6.8	107	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.399	0.3	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	100	-0.02
16	Acenaphthylene	0.400	0.390	2.5	107	-0.01
17	Acenaphthene	0.400	0.383	4.3	104	-0.01
18	Fluorene	0.400	0.387	3.3	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.419	-4.7	100	0.00
21	4-Bromophenyl-phenylether	0.400	0.378	5.5	108	-0.01
22	Hexachlorobenzene	0.400	0.413	-3.2	110	-0.01
23	Atrazine	0.400	0.428	-7.0	124	0.00
24	Pentachlorophenol	0.400	0.394	1.5	115	0.00
25	Phenanthrene	0.400	0.383	4.3	106	0.00
26	Anthracene	0.400	0.369	7.8	106	-0.01
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	105	0.00
28	Fluoranthene	0.400	0.369	7.8	105	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
30	Pyrene	0.400	0.390	2.5	105	0.00
31 S	Terphenyl-d14	0.400	0.385	3.8	105	0.00
32	Benzo(a)anthracene	0.400	0.365	8.8	102	0.00
33	Chrysene	0.400	0.389	2.8	107	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.417	-4.2	117	0.00
35 I	Perylene-d12	0.400	0.400	0.0	95	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	91	0.00
37	Benzo(b)fluoranthene	0.400	0.412	-3.0	110	0.00
38	Benzo(k)fluoranthene	0.400	0.378	5.5	93	0.00
39 C	Benzo(a)pyrene	0.400	0.376	6.0	97	0.00
40	Dibenzo(a,h)anthracene	0.400	0.358	10.5	91	0.00
41	Benzo(g,h,i)perylene	0.400	0.409	-2.2	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4