

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
 Data File : BN021354.D
 Acq On : 25 Aug 2022 02:10
 Operator : CG/JU
 Sample : N4245-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW178I-20220816

Quant Time: Aug 25 07:16:36 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 06:55:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	2773	0.400	ng	0.00
7) Naphthalene-d8	10.920	136	8669	0.400	ng	0.00
13) Acenaphthene-d10	14.739	164	4755	0.400	ng	0.00
19) Phenanthrene-d10	17.490	188	10203	0.400	ng	0.00
29) Chrysene-d12	21.673	240	9245	0.400	ng	# 0.00
35) Perylene-d12	24.200	264	7673	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.606	112	693	0.092	ng	0.00
5) Phenol-d6	7.231	99	527	0.056	ng	0.00
8) Nitrobenzene-d5	9.254	82	1640	0.250	ng	0.00
11) 2-Methylnaphthalene-d10	12.501	152	4306	0.294	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	516	0.237	ng	0.00
15) 2-Fluorobiphenyl	13.371	172	6608	0.311	ng	0.00
27) Fluoranthene-d10	19.514	212	10403	0.341	ng	0.00
31) Terphenyl-d14	20.106	244	8038	0.412	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.411	88	426	0.111	ng	# 83
34) Bis(2-ethylhexyl)phtha...	21.575	149	1469	0.087	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
Data File : BN021354.D
Acq On : 25 Aug 2022 02:10
Operator : CG/JU
Sample : N4245-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178I-20220816

Quant Time: Aug 25 07:16:36 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 06:55:06 2022
Response via : Initial Calibration

