

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
 Data File : BN021361.D
 Acq On : 25 Aug 2022 07:49
 Operator : CG/JU
 Sample : N4245-07
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP02-20220816

Quant Time: Aug 25 08:16:06 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 06:55:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	2769	0.400	ng	0.00
7) Naphthalene-d8	10.920	136	8789	0.400	ng	# 0.00
13) Acenaphthene-d10	14.739	164	4782	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	10156	0.400	ng	#-0.01
29) Chrysene-d12	21.673	240	8782	0.400	ng	# 0.00
35) Perylene-d12	24.203	264	7142	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.606	112	658	0.088	ng	0.00
5) Phenol-d6	7.231	99	500	0.053	ng	0.00
8) Nitrobenzene-d5	9.254	82	1726	0.259	ng	0.00
11) 2-Methylnaphthalene-d10	12.501	152	5359	0.361	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	552	0.252	ng	0.00
15) 2-Fluorobiphenyl	13.371	172	6591	0.308	ng	0.00
27) Fluoranthene-d10	19.514	212	10815	0.356	ng	0.00
31) Terphenyl-d14	20.106	244	7723	0.416	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.411	88	525	0.137	ng	98
6) bis(2-Chloroethyl)ether	7.498	93	267	0.028	ng	94
28) Fluoranthene	19.544	202	891	0.022	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.575	149	1208	0.075	ng	# 97
37) Benzo(b)fluoranthene	23.425	252	729	0.023	ng	# 28

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
Data File : BN021361.D
Acq On : 25 Aug 2022 07:49
Operator : CG/JU
Sample : N4245-07
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP02-20220816

Quant Time: Aug 25 08:16:06 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 06:55:06 2022
Response via : Initial Calibration

