

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
 Data File : BN021366.D
 Acq On : 25 Aug 2022 10:58
 Operator : CG/JU
 Sample : N4245-14
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE126D1-20220817

Quant Time: Aug 25 16:22:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 06:55:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	3261	0.400	ng	0.00
7) Naphthalene-d8	10.920	136	9988	0.400	ng	0.00
13) Acenaphthene-d10	14.739	164	5241	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	11184	0.400	ng	#-0.01
29) Chrysene-d12	21.673	240	9115	0.400	ng	# 0.00
35) Perylene-d12	24.200	264	7298	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.606	112	725	0.082	ng	0.00
5) Phenol-d6	7.231	99	524	0.047	ng	0.00
8) Nitrobenzene-d5	9.254	82	1895	0.251	ng	0.00
11) 2-Methylnaphthalene-d10	12.501	152	5556	0.329	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	526	0.219	ng	0.00
15) 2-Fluorobiphenyl	13.371	172	9171	0.391	ng	0.00
27) Fluoranthene-d10	19.514	212	10683	0.320	ng	0.00
31) Terphenyl-d14	20.106	244	8670	0.450	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.410	88	10714	2.369	ng	# 89
34) Bis(2-ethylhexyl)phtha...	21.575	149	1079	0.065	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082422\
Data File : BN021366.D
Acq On : 25 Aug 2022 10:58
Operator : CG/JU
Sample : N4245-14
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE126D1-20220817

Quant Time: Aug 25 16:22:58 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 06:55:06 2022
Response via : Initial Calibration

