

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011610.D
 Acq On : 25 Aug 2020 10:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 25 13:04:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	145	0.00
2	1,4-Dioxane	0.620	0.636	-2.6	153#	0.00
3	n-Nitrosodimethylamine	0.653	0.719	-10.1	149	0.00
4 S	2-Fluorophenol	1.279	1.236	3.4	138	0.00
5 S	Phenol-d6	1.808	1.736	4.0	132	0.00
6	bis(2-Chloroethyl)ether	1.285	1.332	-3.7	148	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	140	0.00
8 S	Nitrobenzene-d5	0.346	0.357	-3.2	149	-0.01
9	Naphthalene	1.137	1.131	0.5	149	0.00
10	Hexachlorobutadiene	0.238	0.242	-1.7	153#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.757	0.740	2.2	149	0.00
12	2-Methylnaphthalene	0.839	0.824	1.8	147	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	136	0.00
14 S	2,4,6-Tribromophenol	0.220	0.196	10.9	123	0.00
15 S	2-Fluorobiphenyl	1.557	1.555	0.1	142	-0.01
16	Acenaphthylene	2.108	1.972	6.5	139	-0.01
17	Acenaphthene	1.337	1.317	1.5	143	0.00
18	Fluorene	1.730	1.715	0.9	146	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4,6-Dinitro-2-methylphenol	0.046	0.049	-6.5	178#	-0.01
21	4-Bromophenyl-phenylether	0.233	0.229	1.7	148	0.00
22	Hexachlorobenzene	0.270	0.268	0.7	145	0.00
23	Atrazine	0.230	0.160	30.4#	136	0.00
24	Pentachlorophenol	0.119	0.104	12.6	128	0.00
25	Phenanthrene	1.243	1.219	1.9	145	0.00
26	Anthracene	1.224	1.153	5.8	139	0.00
27 SURR	Fluoranthene-d10	1.304	1.264	3.1	149	0.00
28	Fluoranthene	1.506	1.487	1.3	146	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
30	Pyrene	1.418	1.391	1.9	142	0.00
31 S	Terphenyl-d14	1.022	1.044	-2.2	144	0.00
32	Benzo(a)anthracene	1.443	1.442	0.1	142	0.00
33	Chrysene	1.375	1.393	-1.3	151#	0.00
34	Bis(2-ethylhexyl)phthalate	1.027	0.683	33.5#	101	0.00
35	Indeno(1,2,3-cd)pyrene	1.679	1.706	-1.6	146	0.00
36 I	Perylene-d12	1.000	1.000	0.0	139	0.00
37	Benzo(b)fluoranthene	1.449	1.467	-1.2	151#	0.00
38	Benzo(k)fluoranthene	1.441	1.447	-0.4	147	0.00
39 C	Benzo(a)pyrene	1.351	1.323	2.1	144	0.00
40	Dibenzo(a,h)anthracene	1.412	1.412	0.0	147	0.00
41	Benzo(g,h,i)perylene	1.370	1.370	0.0	146	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011610.D
Acq On : 25 Aug 2020 10:13
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 25 13:04:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0