

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011624.D
 Acq On : 25 Aug 2020 20:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 00:35:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.419	-4.7	118	0.00
3	n-Nitrosodimethylamine	0.400	0.421	-5.2	107	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	103	0.00
5 S	Phenol-d6	0.400	0.381	4.8	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.411	-2.7	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.00
8 S	Nitrobenzene-d5	0.400	0.419	-4.7	112	-0.01
9	Naphthalene	0.400	0.399	0.3	110	0.00
10	Hexachlorobutadiene	0.400	0.410	-2.5	113	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	108	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	82	0.00
15 S	2-Fluorobiphenyl	0.400	0.409	-2.2	103	-0.01
16	Acenaphthylene	0.400	0.367	8.3	97	-0.01
17	Acenaphthene	0.400	0.397	0.8	102	0.00
18	Fluorene	0.400	0.386	3.5	101	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.439	-9.7	124	-0.01
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	100	0.00
22	Hexachlorobenzene	0.400	0.412	-3.0	101	0.00
23	Atrazine	0.400	0.268	33.0#	89	0.00
24	Pentachlorophenol	0.400	0.335	16.3	83	0.00
25	Phenanthrene	0.400	0.401	-0.3	100	0.00
26	Anthracene	0.400	0.369	7.8	92	0.00
27 SURR	Fluoranthene-d10	0.400	0.377	5.8	98	0.00
28	Fluoranthene	0.400	0.384	4.0	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.401	-0.3	94	0.00
31 S	Terphenyl-d14	0.400	0.415	-3.7	95	0.00
32	Benzo(a)anthracene	0.400	0.403	-0.8	93	0.00
33	Chrysene	0.400	0.420	-5.0	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.254	36.5#	66	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.422	-5.5	98	0.00
36 I	Perylene-d12	0.400	0.400	0.0	92	0.00
37	Benzo(b)fluoranthene	0.400	0.410	-2.5	101	0.00
38	Benzo(k)fluoranthene	0.400	0.404	-1.0	99	0.00
39 C	Benzo(a)pyrene	0.400	0.395	1.3	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.404	-1.0	99	0.00
41	Benzo(g,h,i)perylene	0.400	0.404	-1.0	98	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011624.D
Acq On : 25 Aug 2020 20:12
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 00:35:51 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0