

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011626.D
 Acq On : 25 Aug 2020 22:00
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Aug 26 00:51:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.620	0.638	-2.9	104	0.00
3	n-Nitrosodimethylamine	0.653	0.747	-14.4	104	0.00
4 S	2-Fluorophenol	1.279	1.177	8.0	89	0.00
5 S	Phenol-d6	1.808	1.631	9.8	84	0.00
6	bis(2-Chloroethyl)ether	1.285	1.298	-1.0	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.346	0.368	-6.4	99	-0.01
9	Naphthalene	1.137	1.130	0.6	96	0.00
10	Hexachlorobutadiene	0.238	0.251	-5.5	102	-0.01
11 SURR	2-Methylnaphthalene-d10	0.757	0.737	2.6	96	0.00
12	2-Methylnaphthalene	0.839	0.827	1.4	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.220	0.190	13.6	75	0.00
15 S	2-Fluorobiphenyl	1.557	1.628	-4.6	94	-0.01
16	Acenaphthylene	2.108	1.941	7.9	86	-0.01
17	Acenaphthene	1.337	1.344	-0.5	92	0.00
18	Fluorene	1.730	1.703	1.6	91	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.046	0.053	-15.2	119	-0.01
21	4-Bromophenyl-phenylether	0.233	0.230	1.3	92	0.00
22	Hexachlorobenzene	0.270	0.280	-3.7	93	0.00
23	Atrazine	0.230	0.154	33.0#	81	0.00
24	Pentachlorophenol	0.119	0.100	16.0	76	0.00
25	Phenanthrene	1.243	1.241	0.2	91	0.00
26	Anthracene	1.224	1.113	9.1	83	0.00
27 SURR	Fluoranthene-d10	1.304	1.209	7.3	88	0.00
28	Fluoranthene	1.506	1.448	3.9	87	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	79	-0.01
30	Pyrene	1.418	1.431	-0.9	85	0.00
31 S	Terphenyl-d14	1.022	1.064	-4.1	85	0.00
32	Benzo(a)anthracene	1.443	1.422	1.5	82	0.00
33	Chrysene	1.375	1.433	-4.2	90	0.00
34	Bis(2-ethylhexyl)phthalate	1.027	0.629	38.8#	54	-0.01
35	Indeno(1,2,3-cd)pyrene	1.679	1.781	-6.1	89	0.00
36 I	Perylene-d12	1.000	1.000	0.0	81	0.00
37	Benzo(b)fluoranthene	1.449	1.481	-2.2	88	0.00
38	Benzo(k)fluoranthene	1.441	1.525	-5.8	90	0.00
39 C	Benzo(a)pyrene	1.351	1.326	1.9	83	0.00
40	Dibenzo(a,h)anthracene	1.412	1.476	-4.5	89	-0.01
41	Benzo(g,h,i)perylene	1.370	1.429	-4.3	88	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011626.D
Acq On : 25 Aug 2020 22:00
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 00:51:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0