

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011626.D
 Acq On : 25 Aug 2020 22:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 00:51:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.411	-2.7	104	0.00
3	n-Nitrosodimethylamine	0.400	0.457	-14.2	104	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	89	0.00
5 S	Phenol-d6	0.400	0.361	9.8	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	90	0.00
8 S	Nitrobenzene-d5	0.400	0.425	-6.2	99	-0.01
9	Naphthalene	0.400	0.397	0.8	96	0.00
10	Hexachlorobutadiene	0.400	0.421	-5.2	102	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	96	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.400	0.345	13.8	75	0.00
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	94	-0.01
16	Acenaphthylene	0.400	0.368	8.0	86	-0.01
17	Acenaphthene	0.400	0.402	-0.5	92	0.00
18	Fluorene	0.400	0.394	1.5	91	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.462	-15.5	119	-0.01
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	92	0.00
22	Hexachlorobenzene	0.400	0.415	-3.7	93	0.00
23	Atrazine	0.400	0.268	33.0#	81	0.00
24	Pentachlorophenol	0.400	0.337	15.8	76	0.00
25	Phenanthrene	0.400	0.399	0.3	91	0.00
26	Anthracene	0.400	0.364	9.0	83	0.00
27 SURR	Fluoranthene-d10	0.400	0.371	7.3	88	0.00
28	Fluoranthene	0.400	0.385	3.8	87	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	79	-0.01
30	Pyrene	0.400	0.404	-1.0	85	0.00
31 S	Terphenyl-d14	0.400	0.416	-4.0	85	0.00
32	Benzo(a)anthracene	0.400	0.394	1.5	82	0.00
33	Chrysene	0.400	0.417	-4.2	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.221	44.8#	54	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.424	-6.0	89	0.00
36 I	Perylene-d12	0.400	0.400	0.0	81	0.00
37	Benzo(b)fluoranthene	0.400	0.409	-2.2	88	0.00
38	Benzo(k)fluoranthene	0.400	0.423	-5.7	90	0.00
39 C	Benzo(a)pyrene	0.400	0.393	1.8	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.418	-4.5	89	-0.01
41	Benzo(g,h,i)perylene	0.400	0.417	-4.2	88	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011626.D
Acq On : 25 Aug 2020 22:00
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 00:51:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0