

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011642.D
 Acq On : 26 Aug 2020 07:38
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Aug 26 08:09:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.620	0.613	1.1	144	0.00
3	n-Nitrosodimethylamine	0.653	0.767	-17.5	155#	0.00
4 S	2-Fluorophenol	1.279	1.149	10.2	125	0.00
5 S	Phenol-d6	1.808	1.598	11.6	119	0.00
6	bis(2-Chloroethyl)ether	1.285	1.298	-1.0	141	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.346	0.372	-7.5	142	-0.01
9	Naphthalene	1.137	1.140	-0.3	138	-0.01
10	Hexachlorobutadiene	0.238	0.256	-7.6	147	-0.01
11 SURR	2-Methylnaphthalene-d10	0.757	0.734	3.0	136	0.00
12	2-Methylnaphthalene	0.839	0.829	1.2	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	-0.01
14 S	2,4,6-Tribromophenol	0.220	0.194	11.8	108	0.00
15 S	2-Fluorobiphenyl	1.557	1.644	-5.6	134	-0.01
16	Acenaphthylene	2.108	1.931	8.4	121	-0.01
17	Acenaphthene	1.337	1.358	-1.6	131	-0.01
18	Fluorene	1.730	1.729	0.1	131	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.01
20	4,6-Dinitro-2-methylphenol	0.046	0.059	-28.3#	186#	-0.01
21	4-Bromophenyl-phenylether	0.233	0.234	-0.4	132	0.00
22	Hexachlorobenzene	0.270	0.283	-4.8	133	0.00
23	Atrazine	0.230	0.149	35.2#	111	0.00
24	Pentachlorophenol	0.119	0.099	16.8	105	0.00
25	Phenanthrene	1.243	1.250	-0.6	129	0.00
26	Anthracene	1.224	1.117	8.7	117	-0.01
27 SURR	Fluoranthene-d10	1.304	1.207	7.4	123	0.00
28	Fluoranthene	1.506	1.466	2.7	125	-0.01
29 I	Chrysene-d12	1.000	1.000	0.0	111	-0.01
30	Pyrene	1.418	1.439	-1.5	120	-0.01
31 S	Terphenyl-d14	1.022	1.073	-5.0	121	0.00
32	Benzo(a)anthracene	1.443	1.407	2.5	114	0.00
33	Chrysene	1.375	1.434	-4.3	127	-0.01
34	Bis(2-ethylhexyl)phthalate	1.027	0.583	43.2#	70	-0.01
35	Indeno(1,2,3-cd)pyrene	1.679	1.807	-7.6	127	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	113	-0.01
37	Benzo(b)fluoranthene	1.449	1.512	-4.3	126	-0.01
38	Benzo(k)fluoranthene	1.441	1.527	-6.0	126	-0.01
39 C	Benzo(a)pyrene	1.351	1.315	2.7	116	-0.01
40	Dibenzo(a,h)anthracene	1.412	1.497	-6.0	127	-0.02
41	Benzo(g,h,i)perylene	1.370	1.441	-5.2	125	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011642.D
Acq On : 26 Aug 2020 07:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 08:09:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0