

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
 Data File : BN011642.D
 Acq On : 26 Aug 2020 07:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 08:09:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	141	0.00
2	1,4-Dioxane	0.400	0.395	1.3	144	0.00
3	n-Nitrosodimethylamine	0.400	0.470	-17.5	155	0.00
4 S	2-Fluorophenol	0.400	0.359	10.3	125	0.00
5 S	Phenol-d6	0.400	0.354	11.5	119	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	141	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.430	-7.5	142	-0.01
9	Naphthalene	0.400	0.401	-0.3	138	-0.01
10	Hexachlorobutadiene	0.400	0.429	-7.2	147	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	136	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	135	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.422	-5.5	134	-0.01
16	Acenaphthylene	0.400	0.366	8.5	121	-0.01
17	Acenaphthene	0.400	0.406	-1.5	131	-0.01
18	Fluorene	0.400	0.400	0.0	131	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	121	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.511	-27.7#	186	-0.01
21	4-Bromophenyl-phenylether	0.400	0.402	-0.5	132	0.00
22	Hexachlorobenzene	0.400	0.420	-5.0	133	0.00
23	Atrazine	0.400	0.259	35.3#	111	0.00
24	Pentachlorophenol	0.400	0.331	17.3	105	0.00
25	Phenanthrene	0.400	0.402	-0.5	129	0.00
26	Anthracene	0.400	0.365	8.8	117	-0.01
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	123	0.00
28	Fluoranthene	0.400	0.389	2.8	125	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
30	Pyrene	0.400	0.406	-1.5	120	-0.01
31 S	Terphenyl-d14	0.400	0.420	-5.0	121	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	114	0.00
33	Chrysene	0.400	0.417	-4.2	127	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.198	50.5#	70	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.431	-7.7	127	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	113	-0.01
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	126	-0.01
38	Benzo(k)fluoranthene	0.400	0.424	-6.0	126	-0.01
39 C	Benzo(a)pyrene	0.400	0.389	2.8	116	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.424	-6.0	127	-0.02
41	Benzo(g,h,i)perylene	0.400	0.421	-5.2	125	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082520\
Data File : BN011642.D
Acq On : 26 Aug 2020 07:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 08:09:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0