




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016022.D
Acq On : 24 Aug 2021 07:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 24 14:47:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 14:46:59 2021
Response via : Initial Calibration






TIC: BN016022.D\data.ms

(70) Pentachlorophenol (C)**16.904min (0.000) 15128.08 ng****response 278639**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.16
264.00	61.60	62.28
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016022.D
Acq On : 24 Aug 2021 07:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 24 14:47:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 14:46:59 2021
Response via : Initial Calibration



TIC: BN016022.D\data.ms

(77) Benzidine

19.504min (0.000) 0.00 ng

response 1337777

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.00
183.00	13.20	11.95
0.00	0.00	0.00