

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
 Data File : BN016027.D
 Acq On : 24 Aug 2021 11:46
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Aug 24 12:41:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 11:43:13 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.882	152	8636	0.400	ng	0.00
7) Naphthalene-d8	10.685	136	44219	0.400	ng	0.00
13) Acenaphthene-d10	14.510	164	23040	0.400	ng	0.00
19) Phenanthrene-d10	17.249	188	42529	0.400	ng	0.00
29) Chrysene-d12	21.440	240	45578	0.400	ng	0.00
35) Perylene-d12	23.840	264	41736	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	16587	0.738	ng	0.00
5) Phenol-d6	7.052	99	23263	0.738	ng	0.00
8) Nitrobenzene-d5	9.038	82	21606	0.494	ng	0.00
11) 2-Methylnaphthalene-d10	12.265	152	54546	0.837	ng	0.00
14) 2,4,6-Tribromophenol	15.994	330	5672	0.690	ng	0.00
15) 2-Fluorobiphenyl	13.140	172	70759	0.676	ng	0.00
27) Fluoranthene-d10	19.286	212	99105	0.803	ng	0.00
31) Terphenyl-d14	19.882	244	88222	0.838	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.392	88	3816	1.040	ng	# 86
3) n-Nitrosodimethylamine	3.751	42	6070	1.005	ng	98
6) bis(2-Chloroethyl)ether	7.310	93	19538	0.742	ng	100
9) Naphthalene	10.723	128	91932	0.752	ng	100
10) Hexachlorobutadiene	11.015	225	20616	0.786	ng	# 99
12) 2-Methylnaphthalene	12.341	142	60339	0.786	ng	100
16) Acenaphthylene	14.231	152	69347	0.634	ng	100
17) Acenaphthene	14.573	154	51556	0.722	ng	99
18) Fluorene	15.550	166	62330	0.722	ng	99
20) 4,6-Dinitro-2-methylph...	15.626	198	2138	0.607	ng	# 89
21) 4-Bromophenyl-phenylether	16.446	248	17332	0.693	ng	99
22) Hexachlorobenzene	16.567	284	23579	0.720	ng	99
23) Atrazine	16.713	200	10706	0.831	ng	98
24) Pentachlorophenol	16.908	266	5291	0.687	ng	99
25) Phenanthrene	17.297	178	95951	0.716	ng	99
26) Anthracene	17.383	178	80687	0.708	ng	100
28) Fluoranthene	19.316	202	112494	0.739	ng	100
30) Pyrene	19.678	202	109980	0.795	ng	100
32) Benzo(a)anthracene	21.429	228	111219	0.766	ng	99
33) Chrysene	21.482	228	117465	0.772	ng	100
34) Bis(2-ethylhexyl)phtha...	21.355	149	27191	0.560	ng	100
36) Indeno(1,2,3-cd)pyrene	26.312	276	126157	0.697	ng	99
37) Benzo(b)fluoranthene	23.108	252	118952	0.751	ng	# 96
38) Benzo(k)fluoranthene	23.156	252	112424	0.750	ng	99
39) Benzo(a)pyrene	23.731	252	103512	0.733	ng	# 93
40) Dibenzo(a,h)anthracene	26.329	278	104942	0.699	ng	99
41) Benzo(g,h,i)perylene	27.075	276	107260	0.711	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016027.D
Acq On : 24 Aug 2021 11:46
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Aug 24 12:41:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 11:43:13 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.882 min Scan# 521
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 8636
Ion Ratio Lower Upper
152 100
150 156.2 126.5 189.7
115 69.3 56.5 84.7



#2
1,4-Dioxane
Concen: 1.040 ng
RT: 3.392 min Scan# 34
Delta R.T. 0.001 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion: 88 Resp: 3816
Ion Ratio Lower Upper
88 100
43 34.7 21.0 31.4#
58 79.2 54.8 82.2









#3
n-Nitrosodimethylamine
Concen: 1.005 ng
RT: 3.751 min Scan# 73
Delta R.T. -0.009 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Instrument :
BNA_N
ClientSampled :
SSTDIC0.8

Tgt Ion: 42 Resp: 6070




Ion	Ratio	Lower	Upper
42	100		
74	119.5	93.8	140.8
44	8.2	7.7	11.5




#4
2-Fluorophenol
Concen: 0.738 ng
RT: 5.457 min Scan# 258
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion: 112 Resp: 16587

Ion	Ratio	Lower	Upper
112	100		
64	65.9	52.5	78.7
63	33.6	26.9	40.3









#5
Phenol-d6
Concen: 0.738 ng
RT: 7.052 min Scan# 431
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 99 Resp: 23263
Ion Ratio Lower Upper
99 100
42 19.4 15.6 23.4
71 37.4 29.8 44.8



#6
bis(2-Chloroethyl)ether
Concen: 0.742 ng
RT: 7.310 min Scan# 459
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion: 93 Resp: 19538
Ion Ratio Lower Upper
93 100
63 77.0 62.0 93.0
95 32.8 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.685 min Scan# 755
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	44219
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	5.3	4.2	6.4
68	4.3	3.4	5.2



#8
Nitrobenzene-d5
Concen: 0.494 ng
RT: 9.038 min Scan# 625
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:	82	Resp:	21606
Ion	Ratio	Lower	Upper
82	100		
128	34.9	27.7	41.5
54	49.4	39.4	59.2









#9
Naphthalene
Concen: 0.752 ng
RT: 10.723 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	128	Resp:	91932
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.9	13.3
127	13.3	10.7	16.1



#10
Hexachlorobutadiene
Concen: 0.786 ng
RT: 11.015 min Scan# 781
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:	225	Resp:	20616
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.837 ng
RT: 12.265 min Scan# 990
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 54546
Ion Ratio Lower Upper
152 100
151 20.5 16.6 24.8



#12
2-Methylnaphthalene
Concen: 0.786 ng
RT: 12.341 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:142 Resp: 60339
Ion Ratio Lower Upper
142 100
141 88.0 70.3 105.5
115 36.0 28.7 43.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.510 min Scan# 1202
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 23040
Ion Ratio Lower Upper
164 100
162 100.2 80.2 120.4
160 44.0 35.9 53.9



#14
2,4,6-Tribromophenol
Concen: 0.690 ng
RT: 15.994 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:330 Resp: 5672
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.9 24.6 36.8









#15
2-Fluorobiphenyl
Concen: 0.676 ng
RT: 13.140 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 70759
Ion Ratio Lower Upper
172 100
171 33.9 27.1 40.7
170 21.4 17.3 25.9



#16
Acenaphthylene
Concen: 0.634 ng
RT: 14.231 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:152 Resp: 69347
Ion Ratio Lower Upper
152 100
151 19.3 15.3 22.9
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 0.722 ng
RT: 14.573 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 51556
Ion Ratio Lower Upper
154 100
153 111.1 89.3 133.9
152 53.6 43.5 65.3



#18
Fluorene
Concen: 0.722 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:166 Resp: 62330
Ion Ratio Lower Upper
166 100
165 97.0 78.2 117.4
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	188	Resp:	42529
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	11.6	9.0	13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.607 ng
RT: 15.626 min Scan# 1290
Delta R.T. -0.012 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:	198	Resp:	2138
Ion Ratio	Lower	Upper	
198	100		
182	15.8	16.8	25.2#
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.693 ng
RT: 16.446 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	248	Resp:	17332
Ion	Ratio	Lower	Upper
248	100		
250	97.7	78.3	117.5
141	61.0	48.0	72.0



#22
Hexachlorobenzene
Concen: 0.720 ng
RT: 16.567 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:	284	Resp:	23579
Ion	Ratio	Lower	Upper
284	100		
142	29.4	23.4	35.2
249	31.5	24.4	36.6









#23
Atrazine
Concen: 0.831 ng
RT: 16.713 min Scan# 1378
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 10706
Ion Ratio Lower Upper
200 100
173 23.3 19.7 29.5
215 50.7 39.5 59.3



#24
Pentachlorophenol
Concen: 0.687 ng
RT: 16.908 min Scan# 1394
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:266 Resp: 5291
Ion Ratio Lower Upper
266 100
264 62.0 50.2 75.4
268 63.8 51.1 76.7









#25
Phenanthrene
Concen: 0.716 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 95951
Ion Ratio Lower Upper
178 100
176 18.2 14.7 22.1
179 15.4 12.6 19.0



#26
Anthracene
Concen: 0.708 ng
RT: 17.383 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:178 Resp: 80687
Ion Ratio Lower Upper
178 100
176 17.5 14.2 21.2
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.803 ng
RT: 19.286 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:212 Resp: 99105
Ion Ratio Lower Upper
212 100
106 11.6 9.4 14.2
104 6.5 5.4 8.0



#28
Fluoranthene
Concen: 0.739 ng
RT: 19.316 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:202 Resp: 112494
Ion Ratio Lower Upper
202 100
101 10.5 8.4 12.6
203 17.5 14.0 21.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.440 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 45578
Ion Ratio Lower Upper
240 100
120 9.6 7.8 11.6
236 26.0 21.0 31.6



#30
Pyrene
Concen: 0.795 ng
RT: 19.678 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:202 Resp: 109980
Ion Ratio Lower Upper
202 100
200 19.8 16.1 24.1
203 17.5 14.1 21.1









#31
Terphenyl-d14
Concen: 0.838 ng
RT: 19.882 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:244 Resp: 88222
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 10.3 8.5 12.7



#32
Benzo(a)anthracene
Concen: 0.766 ng
RT: 21.429 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:228 Resp: 111219
Ion Ratio Lower Upper
228 100
226 25.1 20.6 30.8
229 19.8 16.0 24.0









#33
Chrysene
Concen: 0.772 ng
RT: 21.482 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:228 Resp: 117465
Ion Ratio Lower Upper
228 100
226 27.8 22.2 33.2
229 20.1 16.1 24.1



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.560 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46


Tgt Ion:149 Resp: 27191
Ion Ratio Lower Upper
149 100
167 28.3 22.8 34.2
279 5.2 4.1 6.1








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.840 min Scan# 2531
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:264 Resp: 41736
Ion Ratio Lower Upper
264 100
260 23.8 19.0 28.4
265 28.6 22.7 34.1



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.697 ng
RT: 26.312 min Scan# 3253
Delta R.T. -0.007 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:276 Resp: 126157
Ion Ratio Lower Upper
276 100
138 20.2 16.9 25.3
277 25.3 20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 0.751 ng
RT: 23.108 min Scan# 2317
Delta R.T. -0.003 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 118952
Ion Ratio Lower Upper
252 100
253 22.4 18.6 28.0
125 10.0 10.4 15.6#



#38
Benzo(k)fluoranthene
Concen: 0.750 ng
RT: 23.156 min Scan# 2331
Delta R.T. 0.000 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:252 Resp: 112424
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 8.7 7.5 11.3









#39
Benzo(a)pyrene
Concen: 0.733 ng
RT: 23.731 min Scan# 2499
Delta R.T. -0.003 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 103512
Ion Ratio Lower Upper
252 100
253 22.7 19.2 28.8
125 12.1 14.2 21.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.699 ng
RT: 26.329 min Scan# 3258
Delta R.T. -0.010 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Tgt Ion:278 Resp: 104942
Ion Ratio Lower Upper
278 100
139 13.2 10.7 16.1
279 23.6 19.7 29.5





#41
Benzo(g,h,i)perylene
Concen: 0.711 ng
RT: 27.075 min Scan# 3476
Delta R.T. -0.003 min
Lab File: BN016027.D
Acq: 24 Aug 2021 11:46

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:276 Resp: 107260
Ion Ratio Lower Upper
276 100
277 23.8 19.3 28.9
138 17.2 14.0 21.0

