

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
 Data File : BN016030.D
 Acq On : 24 Aug 2021 13:36
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Aug 24 14:05:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 11:43:13 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.882	152	9931	0.400	ng	0.00
7) Naphthalene-d8	10.685	136	46176	0.400	ng	0.00
13) Acenaphthene-d10	14.510	164	25293	0.400	ng	0.00
19) Phenanthrene-d10	17.249	188	47596	0.400	ng	0.00
29) Chrysene-d12	21.450	240	49098	0.400	ng	# 0.01
35) Perylene-d12	23.844	264	46820	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	117172	4.536	ng	0.00
5) Phenol-d6	7.052	99	165724	4.572	ng	0.00
8) Nitrobenzene-d5	9.038	82	192166	4.206	ng	0.00
11) 2-Methylnaphthalene-d10	12.265	152	349142	5.128	ng	0.00
14) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
15) 2-Fluorobiphenyl	13.140	172	439895	3.828	ng	0.00
27) Fluoranthene-d10	19.286	212	711029	5.145	ng	0.00
31) Terphenyl-d14	19.882	244	588950	5.193	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.392	88	25760	6.107	ng	92
3) n-Nitrosodimethylamine	3.724	42	48911	7.042	ng	# 93
6) bis(2-Chloroethyl)ether	7.310	93	124853	4.125	ng	99
9) Naphthalene	10.724	128	568487	4.451	ng	99
10) Hexachlorobutadiene	11.015	225	129897	4.743	ng	# 99
12) 2-Methylnaphthalene	12.341	142	369742	4.614	ng	99
16) Acenaphthylene	14.231	152	531301	4.425	ng	99
17) Acenaphthene	14.573	154	343463	4.384	ng	96
18) Fluorene	15.550	166	417225	4.399	ng	99
21) 4-Bromophenyl-phenylether	16.446	248	120894	4.318	ng	99
22) Hexachlorobenzene	16.567	284	155514	4.243	ng	99
24) Pentachlorophenol	16.908	266	66287	7.692	ng	100
25) Phenanthrene	17.297	178	633814	4.226	ng	99
26) Anthracene	17.383	178	602786	4.723	ng	99
28) Fluoranthene	19.316	202	727798	4.273	ng	100
30) Pyrene	19.678	202	747557	5.016	ng	100
32) Benzo(a)anthracene	21.429	228	784414	5.017	ng	100
33) Chrysene	21.482	228	735822	4.491	ng	98
36) Indeno(1,2,3-cd)pyrene	26.325	276	844613	4.162	ng	99
37) Benzo(b)fluoranthene	23.115	252	786828	4.427	ng	# 95
38) Benzo(k)fluoranthene	23.163	252	721212	4.289	ng	98
39) Benzo(a)pyrene	23.738	252	710953	4.488	ng	# 90
40) Dibenzo(a,h)anthracene	26.343	278	704154	4.180	ng	98
41) Benzo(g,h,i)perylene	27.089	276	730885	4.317	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016030.D
Acq On : 24 Aug 2021 13:36
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Aug 24 14:05:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 11:43:13 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.882 min Scan# 521
 Delta R.T. 0.000 min
 Lab File: BN016030.D
 Acq: 24 Aug 2021 13:36

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:152 Resp: 9931




Ion	Ratio	Lower	Upper
152	100		
150	156.8	126.5	189.7
115	70.8	56.5	84.7




#2
 1,4-Dioxane
 Concen: 6.107 ng
 RT: 3.392 min Scan# 34
 Delta R.T. 0.001 min
 Lab File: BN016030.D
 Acq: 24 Aug 2021 13:36

Tgt Ion: 88 Resp: 25760

Ion	Ratio	Lower	Upper
88	100		
43	26.1	21.0	31.4
58	77.9	54.8	82.2









#3
n-Nitrosodimethylamine
Concen: 7.042 ng
RT: 3.724 min Scan# 70
Delta R.T. -0.037 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 48911
Ion Ratio Lower Upper
42 100
74 125.3 93.8 140.8
44 6.6 7.7 11.5#



#4
2-Fluorophenol
Concen: 4.536 ng
RT: 5.457 min Scan# 258
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion: 112 Resp: 117172
Ion Ratio Lower Upper
112 100
64 65.9 52.5 78.7
63 33.9 26.9 40.3









#5
Phenol-d6
Concen: 4.572 ng
RT: 7.052 min Scan# 431
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 165724
Ion Ratio Lower Upper
99 100
42 19.4 15.6 23.4
71 37.9 29.8 44.8



#6
bis(2-Chloroethyl)ether
Concen: 4.125 ng
RT: 7.310 min Scan# 459
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion: 93 Resp: 124853
Ion Ratio Lower Upper
93 100
63 76.0 62.0 93.0
95 32.4 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.685 min Scan# 755
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	46176
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	5.2	4.2	6.4
68	4.3	3.4	5.2



#8
Nitrobenzene-d5
Concen: 4.206 ng
RT: 9.038 min Scan# 625
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:	82	Resp:	192166
Ion	Ratio	Lower	Upper
82	100		
128	37.5	27.7	41.5
54	46.6	39.4	59.2









#9
Naphthalene
Concen: 4.451 ng
RT: 10.724 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 568487
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.2 10.7 16.1



#10
Hexachlorobutadiene
Concen: 4.743 ng
RT: 11.015 min Scan# 781
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:225 Resp: 129897
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 5.128 ng
RT: 12.265 min Scan# 990
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 349142
Ion Ratio Lower Upper
152 100
151 20.2 16.6 24.8



#12
2-Methylnaphthalene
Concen: 4.614 ng
RT: 12.341 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:142 Resp: 369742
Ion Ratio Lower Upper
142 100
141 87.2 70.3 105.5
115 35.4 28.7 43.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.510 min Scan# 1202
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 25293
Ion Ratio Lower Upper
164 100
162 99.0 80.2 120.4
160 43.2 35.9 53.9




#14
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.99 min

Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion: 330
Sig Exp Ratio
330 100
332 0.0
141 30.7









#15
2-Fluorobiphenyl
Concen: 3.828 ng
RT: 13.140 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:172 Resp: 439895




Ion	Ratio	Lower	Upper
172	100		
171	34.0	27.1	40.7
170	21.5	17.3	25.9




#16
Acenaphthylene
Concen: 4.425 ng
RT: 14.231 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:152 Resp: 531301

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.3	22.9
153	13.2	10.5	15.7









#17
Acenaphthene
Concen: 4.384 ng
RT: 14.573 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 343463
Ion Ratio Lower Upper
154 100
153 108.7 89.3 133.9
152 50.7 43.5 65.3



#18
Fluorene
Concen: 4.399 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:166 Resp: 417225
Ion Ratio Lower Upper
166 100
165 96.5 78.2 117.4
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188	Resp:	47596
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	11.5	9.0



#21
4-Bromophenyl-phenylether
Concen: 4.318 ng
RT: 16.446 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:248	Resp:	120894
Ion Ratio	Lower	Upper
248	100	
250	97.8	78.3
141	61.2	48.0









#22
Hexachlorobenzene
Concen: 4.243 ng
RT: 16.567 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 155514
Ion Ratio Lower Upper
284 100
142 29.6 23.4 35.2
249 31.2 24.4 36.6



#24
Pentachlorophenol
Concen: 7.692 ng
RT: 16.908 min Scan# 1394
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:266 Resp: 66287
Ion Ratio Lower Upper
266 100
264 62.4 50.2 75.4
268 64.3 51.1 76.7









#25
Phenanthrene
Concen: 4.226 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 633814
Ion Ratio Lower Upper
178 100
176 18.1 14.7 22.1
179 15.4 12.6 19.0



#26
Anthracene
Concen: 4.723 ng
RT: 17.383 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:178 Resp: 602786
Ion Ratio Lower Upper
178 100
176 17.7 14.2 21.2
179 15.7 12.2 18.2









#27
Fluoranthene-d10
Concen: 5.145 ng
RT: 19.286 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 711029
Ion Ratio Lower Upper
212 100
106 11.5 9.4 14.2
104 6.5 5.4 8.0



#28
Fluoranthene
Concen: 4.273 ng
RT: 19.316 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:202 Resp: 727798
Ion Ratio Lower Upper
202 100
101 10.5 8.4 12.6
203 17.6 14.0 21.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.450 min Scan# 2022
Delta R.T. 0.011 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 49098
Ion Ratio Lower Upper
240 100
120 5.5 7.8 11.6#
236 24.9 21.0 31.6



#30
Pyrene
Concen: 5.016 ng
RT: 19.678 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:202 Resp: 747557
Ion Ratio Lower Upper
202 100
200 19.9 16.1 24.1
203 17.6 14.1 21.1









#31
 Terphenyl-d14
 Concen: 5.193 ng
 RT: 19.882 min Scan# 1861
 Delta R.T. 0.000 min
 Lab File: BN016030.D
 Acq: 24 Aug 2021 13:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 588950
 Ion Ratio Lower Upper
 244 100
 212 7.2 6.0 9.0
 122 10.6 8.5 12.7



#32
 Benzo(a)anthracene
 Concen: 5.017 ng
 RT: 21.429 min Scan# 2020
 Delta R.T. 0.000 min
 Lab File: BN016030.D
 Acq: 24 Aug 2021 13:36

Tgt Ion:228 Resp: 784414
 Ion Ratio Lower Upper
 228 100
 226 25.7 20.6 30.8
 229 19.8 16.0 24.0









#33
Chrysene
Concen: 4.491 ng
RT: 21.482 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 735822
Ion Ratio Lower Upper
228 100
226 28.8 22.2 33.2
229 19.5 16.1 24.1



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.844 min Scan# 2532
Delta R.T. 0.004 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:264 Resp: 46820
Ion Ratio Lower Upper
264 100
260 23.8 19.0 28.4
265 27.1 22.7 34.1









#36
Indeno(1,2,3-cd)pyrene
Concen: 4.162 ng
RT: 26.325 min Scan# 3257
Delta R.T. 0.007 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 844613
Ion Ratio Lower Upper
276 100
138 19.6 16.9 25.3
277 25.3 20.2 30.4



#37
Benzo(b)fluoranthene
Concen: 4.427 ng
RT: 23.115 min Scan# 2319
Delta R.T. 0.004 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:252 Resp: 786828
Ion Ratio Lower Upper
252 100
253 22.2 18.6 28.0
125 8.5 10.4 15.6#









#38
Benzo(k)fluoranthene
Concen: 4.289 ng
RT: 23.163 min Scan# 2333
Delta R.T. 0.007 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 721212
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.0 7.5 11.3



#39
Benzo(a)pyrene
Concen: 4.488 ng
RT: 23.738 min Scan# 2501
Delta R.T. 0.004 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:252 Resp: 710953
Ion Ratio Lower Upper
252 100
253 22.0 19.2 28.8
125 9.4 14.2 21.2#









#40
Dibenzo(a,h)anthracene
Concen: 4.180 ng
RT: 26.343 min Scan# 3262
Delta R.T. 0.004 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 704154
Ion Ratio Lower Upper
278 100
139 12.5 10.7 16.1
279 23.6 19.7 29.5



#41
Benzo(g,h,i)perylene
Concen: 4.317 ng
RT: 27.089 min Scan# 3480
Delta R.T. 0.011 min
Lab File: BN016030.D
Acq: 24 Aug 2021 13:36

Tgt Ion:276 Resp: 730885
Ion Ratio Lower Upper
276 100
277 23.6 19.3 28.9
138 16.9 14.0 21.0

