



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
 Data File : BN016022.D
 Acq On : 24 Aug 2021 07:05
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082521.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 24 14:26:02 2021






AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.5	151437	PASS
68	69	0.00	2	1.9	3079	PASS
69	198	0.00	100	44.7	159088	PASS
70	69	0.00	2	0.6	996	PASS
127	198	10	80	48.9	174019	PASS
197	198	0.00	2	0.5	1742	PASS
198	198	100	100	100.0	356224	PASS
199	198	5	9	6.6	23619	PASS
275	198	10	60	28.3	100965	PASS
365	198	1	100	4.2	14969	PASS
441	198	0.01	100	14.4	51325	PASS
442	442	50	100	100.0	323733	PASS
443	442	15	24	19.1	61853	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016022.D
Acq On : 24 Aug 2021 07:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 24 14:47:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 14:46:59 2021
Response via : Initial Calibration



TIC: BN016022.D\data.ms

(70) Pentachlorophenol (C)

16.904min (0.000) 15128.08 ng




response 278639

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.16
264.00	61.60	62.28
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082521\
Data File : BN016022.D
Acq On : 24 Aug 2021 07:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 24 14:47:05 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 14:46:59 2021
Response via : Initial Calibration



TIC: BN016022.D\data.ms

(77) Benzidine

19.504min (0.000) 0.00 ng

response 1337777

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.00
183.00	13.20	11.95
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2021	BNA_N	<u>BN016022.D</u>
Compound Name	Response	Retention Time
DDT	1090013	20.739
DDD	3581	20.298
DDE	1083	19.792
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4664	1094677	0.43

Instrument :
BNA_N
ClientSampleId :
DFTPP