

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
 Data File : BN021388.D
 Acq On : 26 Aug 2022 16:57
 Operator : CG/JU
 Sample : N4245-17
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT163S1-20220817

Quant Time: Aug 26 23:41:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:37:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	4087	0.400	ng	0.00
7) Naphthalene-d8	10.909	136	12388	0.400	ng	0.00
13) Acenaphthene-d10	14.728	164	6441	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	13622	0.400	ng	# 0.00
29) Chrysene-d12	21.673	240	9975	0.400	ng	# 0.00
35) Perylene-d12	24.202	264	7349	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.606	112	988	0.089	ng	0.00
5) Phenol-d6	7.231	99	730	0.053	ng	0.00
8) Nitrobenzene-d5	9.254	82	2446	0.261	ng	0.00
11) 2-Methylnaphthalene-d10	12.497	152	7587	0.363	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	650	0.220	ng	0.01
15) 2-Fluorobiphenyl	13.360	172	11070	0.384	ng	0.00
27) Fluoranthene-d10	19.514	212	13633	0.335	ng	0.00
31) Terphenyl-d14	20.106	244	11396	0.541	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.410	88	1712	0.302	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.574	149	1047	0.058	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
Data File : BN021388.D
Acq On : 26 Aug 2022 16:57
Operator : CG/JU
Sample : N4245-17
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT163S1-20220817

Quant Time: Aug 26 23:41:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 26 23:37:03 2022
Response via : Initial Calibration

