

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
 Data File : BN021389.D
 Acq On : 26 Aug 2022 17:36
 Operator : CG/JU
 Sample : N4245-19
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D3-20220817

Quant Time: Aug 26 23:42:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:37:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	4164	0.400	ng	0.00
7) Naphthalene-d8	10.919	136	12622	0.400	ng	# 0.01
13) Acenaphthene-d10	14.728	164	6691	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	13761	0.400	ng	# 0.00
29) Chrysene-d12	21.673	240	10065	0.400	ng	# 0.00
35) Perylene-d12	24.200	264	7968	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.606	112	1069	0.095	ng	0.00
5) Phenol-d6	7.231	99	803	0.057	ng	0.00
8) Nitrobenzene-d5	9.254	82	2667	0.279	ng	0.00
11) 2-Methylnaphthalene-d10	12.497	152	8308	0.390	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	578	0.189	ng	0.01
15) 2-Fluorobiphenyl	13.360	172	10876	0.363	ng	0.00
27) Fluoranthene-d10	19.514	212	14580	0.355	ng	0.00
31) Terphenyl-d14	20.106	244	10208	0.480	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.410	88	16781	2.905	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.574	149	1090	0.059	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
Data File : BN021389.D
Acq On : 26 Aug 2022 17:36
Operator : CG/JU
Sample : N4245-19
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D3-20220817

Quant Time: Aug 26 23:42:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 26 23:37:03 2022
Response via : Initial Calibration

