

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
 Data File : BN021381.D
 Acq On : 26 Aug 2022 12:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 23:38:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:37:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	165	0.00
2	1,4-Dioxane	0.400	0.385	3.8	155	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	145	0.00
4 S	2-Fluorophenol	0.400	0.313	21.8	128	0.00
5 S	Phenol-d6	0.400	0.310	22.5	127	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	153	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	157	0.00
8 S	Nitrobenzene-d5	0.400	0.342	14.5	143	0.00
9	Naphthalene	0.400	0.351	12.3	146	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	158	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.467	-16.8	190	0.00
12	2-Methylnaphthalene	0.400	0.264	34.0#	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.400	0.269	32.8#	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.395	1.3	146	0.00
16	Acenaphthylene	0.400	0.354	11.5	139	0.00
17	Acenaphthene	0.400	0.369	7.8	140	0.00
18	Fluorene	0.400	0.361	9.8	136	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.054	86.5#	25	0.00
21	4-Bromophenyl-phenylether	0.400	0.363	9.3	133	0.00
22	Hexachlorobenzene	0.400	0.386	3.5	139	0.00
23	Atrazine	0.400	0.285	28.8#	109	0.00
24	Pentachlorophenol	0.400	0.218	45.5#	88	0.00
25	Phenanthrene	0.400	0.362	9.5	133	0.00
26	Anthracene	0.400	0.340	15.0	129	0.00
27 SURR	Fluoranthene-d10	0.400	0.339	15.3	125	0.00
28	Fluoranthene	0.400	0.341	14.8	127	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
30	Pyrene	0.400	0.407	-1.7	123	0.00
31 S	Terphenyl-d14	0.400	0.421	-5.2	129	0.00
32	Benzo(a)anthracene	0.400	0.342	14.5	108	0.00
33	Chrysene	0.400	0.402	-0.5	123	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.279	30.3#	94	0.00
35 I	Perylene-d12	0.400	0.400	0.0	91	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.337	15.8	83	0.00
37	Benzo(b)fluoranthene	0.400	0.416	-4.0	104	0.00
38	Benzo(k)fluoranthene	0.400	0.423	-5.7	102	0.00
39 C	Benzo(a)pyrene	0.400	0.360	10.0	91	0.00
40	Dibenzo(a,h)anthracene	0.400	0.336	16.0	82	0.00
41	Benzo(g,h,i)perylene	0.400	0.353	11.8	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4