

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082622\
 Data File : BN021398.D
 Acq On : 27 Aug 2022 00:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 27 04:06:30 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:37:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	170	0.00
2	1,4-Dioxane	0.400	0.391	2.3	161	0.00
3	n-Nitrosodimethylamine	0.400	0.343	14.2	146	0.00
4 S	2-Fluorophenol	0.400	0.312	22.0	131	0.00
5 S	Phenol-d6	0.400	0.313	21.8	132	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	157	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	164	0.00
8 S	Nitrobenzene-d5	0.400	0.337	15.8	147	0.00
9	Naphthalene	0.400	0.352	12.0	154	0.00
10	Hexachlorobutadiene	0.400	0.384	4.0	164	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.519	-29.8#	220	0.00
12	2-Methylnaphthalene	0.400	0.272	32.0#	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	152	0.00
14 S	2,4,6-Tribromophenol	0.400	0.284	29.0#	123	0.00
15 S	2-Fluorobiphenyl	0.400	0.395	1.3	157	0.00
16	Acenaphthylene	0.400	0.351	12.3	148	0.00
17	Acenaphthene	0.400	0.372	7.0	152	0.00
18	Fluorene	0.400	0.368	8.0	150	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	147	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.049	87.8#	25	-0.03
21	4-Bromophenyl-phenylether	0.400	0.369	7.8	147	0.00
22	Hexachlorobenzene	0.400	0.395	1.3	154	0.00
23	Atrazine	0.400	0.291	27.3#	121	0.00
24	Pentachlorophenol	0.400	0.254	36.5#	112	0.00
25	Phenanthrene	0.400	0.371	7.3	148	0.00
26	Anthracene	0.400	0.344	14.0	142	0.00
27 SURR	Fluoranthene-d10	0.400	0.342	14.5	137	0.00
28	Fluoranthene	0.400	0.350	12.5	142	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	127	0.00
30	Pyrene	0.400	0.410	-2.5	135	0.00
31 S	Terphenyl-d14	0.400	0.428	-7.0	144	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	121	0.00
33	Chrysene	0.400	0.392	2.0	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.277	30.8#	102	0.00
35 I	Perylene-d12	0.400	0.400	0.0	103	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.348	13.0	97	0.00
37	Benzo(b)fluoranthene	0.400	0.412	-3.0	116	-0.01
38	Benzo(k)fluoranthene	0.400	0.396	1.0	107	0.00
39 C	Benzo(a)pyrene	0.400	0.380	5.0	108	0.00
40	Dibenzo(a,h)anthracene	0.400	0.348	13.0	96	0.00
41	Benzo(g,h,i)perylene	0.400	0.347	13.3	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4