

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082922\
 Data File : BN021421.D
 Acq On : 29 Aug 2022 13:42
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Aug 29 14:54:48 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 14:52:01 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.076	152	5351	0.400	ng	0.00
7) Naphthalene-d8	10.909	136	16249	0.400	ng	0.00
13) Acenaphthene-d10	14.728	164	8964	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	17529	0.400	ng	0.00
29) Chrysene-d12	21.673	240	14611	0.400	ng	0.00
35) Perylene-d12	24.197	264	8915	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.599	112	16500	1.139	ng	0.00
5) Phenol-d6	7.224	99	21350	1.176	ng	0.00
8) Nitrobenzene-d5	9.254	82	17656	1.436	ng	0.00
11) 2-Methylnaphthalene-d10	12.493	152	59713	2.176	ng	0.00
14) 2,4,6-Tribromophenol	16.218	330	4700	1.145	ng	0.00
15) 2-Fluorobiphenyl	13.360	172	61329	1.530	ng	0.00
27) Fluoranthene-d10	19.510	212	73600	1.405	ng	0.00
31) Terphenyl-d14	20.097	244	50650	1.641	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.410	88	10589	1.427	ng	# 89
3) n-Nitrosodimethylamine	3.743	42	10203	1.437	ng	# 94
6) bis(2-Chloroethyl)ether	7.491	93	26487	1.459	ng	98
9) Naphthalene	10.962	128	76355	1.429	ng	98
10) Hexachlorobutadiene	11.240	225	13092	1.486	ng	# 100
12) 2-Methylnaphthalene	12.187	142	10978	1.207	ng	# 87
16) Acenaphthylene	14.450	152	70115	1.560	ng	100
17) Acenaphthene	14.793	154	47615	1.527	ng	99
18) Fluorene	15.776	166	59239	1.516	ng	99
20) 4,6-Dinitro-2-methylph...	15.648	198	1	0.043	ng	86
21) 4-Bromophenyl-phenylether	16.670	248	17850	1.487	ng	99
22) Hexachlorobenzene	16.783	284	21240	1.525	ng	99
23) Atrazine	16.926	200	10552	1.391	ng	97
24) Pentachlorophenol	17.121	266	5163	1.073	ng	99
25) Phenanthrene	17.521	178	97766	1.525	ng	100
26) Anthracene	17.613	178	80278	1.524	ng	99
28) Fluoranthene	19.539	202	105994	1.502	ng	100
30) Pyrene	19.902	202	117618	1.599	ng	# 94
32) Benzo(a)anthracene	21.655	228	81409	1.472	ng	99
33) Chrysene	21.709	228	94056	1.560	ng	99
34) Bis(2-ethylhexyl)phtha...	21.565	149	28839	1.082	ng	98
36) Indeno(1,2,3-cd)pyrene	26.857	276	65802	1.451	ng	100
37) Benzo(b)fluoranthene	23.419	252	71795	1.826	ng	# 94
38) Benzo(k)fluoranthene	23.469	252	75419	1.856	ng	95
39) Benzo(a)pyrene	24.083	252	50914	1.596	ng	# 91
40) Dibenzo(a,h)anthracene	26.881	278	54042	1.449	ng	94
41) Benzo(g,h,i)perylene	27.676	276	56853	1.448	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082922\
Data File : BN021421.D
Acq On : 29 Aug 2022 13:42
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Quant Time: Aug 29 14:54:48 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 29 14:52:01 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.076 min Scan# 665
Delta R.T. -0.007 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:152 Resp: 5351
Ion Ratio Lower Upper
152 100
150 157.4 121.4 182.0
115 61.9 48.6 72.8




#2
1,4-Dioxane
Concen: 1.427 ng
RT: 3.410 min Scan# 18
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42



Tgt Ion: 88 Resp: 10589
Ion Ratio Lower Upper
88 100
43 27.7 35.8 53.8#
58 72.6 57.1 85.7









#3
 n-Nitrosodimethylamine
 Concen: 1.437 ng
 RT: 3.743 min Scan# 64
 Delta R.T. -0.007 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion: 42 Resp: 10203
 Ion Ratio Lower Upper
 42 100
 74 134.4 113.0 169.6
 44 7.1 10.5 15.7#



#4
 2-Fluorophenol
 Concen: 1.139 ng
 RT: 5.599 min Scan# 321
 Delta R.T. -0.000 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42

Tgt Ion: 112 Resp: 16500
 Ion Ratio Lower Upper
 112 100
 64 55.9 43.5 65.3
 63 29.0 22.6 34.0









#5
Phenol-d6
Concen: 1.176 ng
RT: 7.224 min Scan# 54
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 21350
Ion Ratio Lower Upper
99 100
42 18.5 15.5 23.3
71 31.2 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 1.459 ng
RT: 7.491 min Scan# 583
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion: 93 Resp: 26487
Ion Ratio Lower Upper
93 100
63 75.1 57.8 86.8
95 32.5 25.9 38.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.909 min Scan# 94
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	16249
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.3	13.9
54	7.8	6.9	10.3
68	6.8	6.1	9.1



#8
Nitrobenzene-d5
Concen: 1.436 ng
RT: 9.254 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:	82	Resp:	17656
Ion	Ratio	Lower	Upper
82	100		
128	45.3	30.9	46.3
54	43.5	42.7	64.1









#9
Naphthalene
Concen: 1.429 ng
RT: 10.962 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 76355
Ion Ratio Lower Upper
128 100
129 11.0 9.3 13.9
127 12.9 11.0 16.4



#10
Hexachlorobutadiene
Concen: 1.486 ng
RT: 11.240 min Scan# 971
Delta R.T. -0.011 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:225 Resp: 13092
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 2.176 ng
RT: 12.493 min Scan# 1176
Delta R.T. -0.004 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 59713
Ion Ratio Lower Upper
152 100
151 19.6 14.5 21.7



#12
2-Methylnaphthalene
Concen: 1.207 ng
RT: 12.187 min Scan# 1176
Delta R.T. -0.004 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:142 Resp: 10978
Ion Ratio Lower Upper
142 100
141 11.2 12.2 18.2#
115 2.8 7.9 11.9#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.728 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 8964
Ion Ratio Lower Upper
164 100
162 103.3 82.2 123.4
160 49.9 39.5 59.3



#14
2,4,6-Tribromophenol
Concen: 1.145 ng
RT: 16.218 min Scan# 1609
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:330 Resp: 4700
Ion Ratio Lower Upper
330 100
332 96.9 79.3 118.9
141 44.6 37.1 55.7









#15
2-Fluorobiphenyl
Concen: 1.530 ng
RT: 13.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 61329
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 23.0 18.6 28.0



#16
Acenaphthylene
Concen: 1.560 ng
RT: 14.450 min Scan# 1442
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:152 Resp: 70115
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.2
153 13.0 10.3 15.5









#17
Acenaphthene
Concen: 1.527 ng
RT: 14.793 min Scan# 1474
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 47615
Ion Ratio Lower Upper
154 100
153 111.0 89.5 134.3
152 56.6 45.5 68.3



#18
Fluorene
Concen: 1.516 ng
RT: 15.776 min Scan# 1566
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:166 Resp: 59239
Ion Ratio Lower Upper
166 100
165 97.6 78.7 118.1
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.480 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 17529
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.5 9.5 14.3



#20
4,6-Dinitro-2-methylphenol
Concen: 0.043 ng
RT: 15.648 min Scan# 1554
Delta R.T. 0.021 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:198 Resp: 1
Ion Ratio Lower Upper
198 100
182 94.0 86.6 130.0
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.487 ng
RT: 16.670 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 17850
Ion Ratio Lower Upper
248 100
250 102.1 81.8 122.6
141 46.2 38.5 57.7



#22
Hexachlorobenzene
Concen: 1.525 ng
RT: 16.783 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:284 Resp: 21240
Ion Ratio Lower Upper
284 100
142 41.0 32.3 48.5
249 31.0 25.4 38.2









#23
 Atrazine
 Concen: 1.391 ng
 RT: 16.926 min Scan# 1678
 Delta R.T. -0.000 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 10552
 Ion Ratio Lower Upper
 200 100
 173 30.1 24.3 36.5
 215 44.5 38.4 57.6



#24
 Pentachlorophenol
 Concen: 1.073 ng
 RT: 17.121 min Scan# 1697
 Delta R.T. -0.010 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42

Tgt Ion:266 Resp: 5163
 Ion Ratio Lower Upper
 266 100
 264 62.9 49.0 73.6
 268 64.2 50.9 76.3









#25
Phenanthrene
Concen: 1.525 ng
RT: 17.521 min Scan# 1736
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 97766
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.2 12.2 18.4



#26
Anthracene
Concen: 1.524 ng
RT: 17.613 min Scan# 1745
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:178 Resp: 80278
Ion Ratio Lower Upper
178 100
176 18.7 14.6 22.0
179 15.1 12.2 18.2









#27
Fluoranthene-d10
Concen: 1.405 ng
RT: 19.510 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 73600
Ion Ratio Lower Upper
212 100
106 18.0 14.2 21.2
104 10.1 7.8 11.6



#28
Fluoranthene
Concen: 1.502 ng
RT: 19.539 min Scan# 2146
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:202 Resp: 105994
Ion Ratio Lower Upper
202 100
101 13.5 10.9 16.3
203 17.1 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.673 min Scan# 2443

Delta R.T. 0.009 min

Lab File: BN021421.D

Acq: 29 Aug 2022 13:42

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 14611

Ion Ratio Lower Upper

240 100

120 12.3 10.2 15.4

236 27.2 22.1 33.1



#30

Pyrene

Concen: 1.599 ng

RT: 19.902 min Scan# 2232

Delta R.T. -0.000 min

Lab File: BN021421.D

Acq: 29 Aug 2022 13:42




Tgt Ion:202 Resp: 117618


Ion Ratio Lower Upper

202 100

200 21.1 13.9 20.9#

203 16.5 11.9 17.9









#31
 Terphenyl-d14
 Concen: 1.641 ng
 RT: 20.097 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 50650
 Ion Ratio Lower Upper
 244 100
 212 9.8 6.5 9.7#
 122 20.7 9.1 13.7#



#32
 Benzo(a)anthracene
 Concen: 1.472 ng
 RT: 21.655 min Scan# 2441
 Delta R.T. -0.000 min
 Lab File: BN021421.D
 Acq: 29 Aug 2022 13:42

Tgt Ion:228 Resp: 81409
 Ion Ratio Lower Upper
 228 100
 226 27.1 22.0 33.0
 229 19.5 15.9 23.9









#33
Chrysene
Concen: 1.560 ng
RT: 21.709 min Scan# 2447
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 94056
Ion Ratio Lower Upper
228 100
226 30.1 24.2 36.2
229 19.4 15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.082 ng
RT: 21.565 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:149 Resp: 28839
Ion Ratio Lower Upper
149 100
167 25.7 21.4 32.0
279 2.5 2.2 3.2









#35
Perylene-d12
Concen: 0.400 ng
RT: 24.197 min Scan# 3136
Delta R.T. 0.003 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264 Resp: 8915
Ion Ratio Lower Upper
264 100
260 26.4 21.0 31.6
265 37.6 32.2 48.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.451 ng
RT: 26.857 min Scan# 4047
Delta R.T. 0.009 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:276 Resp: 65802
Ion Ratio Lower Upper
276 100
138 33.9 27.4 41.2
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 1.826 ng
RT: 23.419 min Scan# 2869
Delta R.T. 0.006 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 71795
Ion Ratio Lower Upper
252 100
253 22.3 19.8 29.6
125 13.5 13.7 20.5#



#38
Benzo(k)fluoranthene
Concen: 1.856 ng
RT: 23.469 min Scan# 2888
Delta R.T. 0.003 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42


Tgt Ion:252 Resp: 75419
Ion Ratio Lower Upper
252 100
253 21.8 19.5 29.3
125 13.6 12.3 18.5








#39
Benzo(a)pyrene
Concen: 1.596 ng
RT: 24.083 min Scan# 3096
Delta R.T. 0.006 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 50914
Ion Ratio Lower Upper
252 100
253 22.6 21.1 31.7
125 15.4 16.6 25.0#



#40
Dibenzo(a,h)anthracene
Concen: 1.449 ng
RT: 26.881 min Scan# 4055
Delta R.T. 0.006 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Tgt Ion:278 Resp: 54042
Ion Ratio Lower Upper
278 100
139 22.9 21.8 32.6
279 23.6 20.4 30.6





#41
Benzo(g,h,i)perylene
Concen: 1.448 ng
RT: 27.676 min Scan# 41
Delta R.T. 0.006 min
Lab File: BN021421.D
Acq: 29 Aug 2022 13:42

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 56853
Ion Ratio Lower Upper
276 100
277 23.4 19.7 29.5
138 28.9 23.7 35.5

