

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082922\
 Data File : BN021424.D
 Acq On : 29 Aug 2022 15:37
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082922

Quant Time: Aug 29 16:42:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.385	3.8	87	0.00
3	n-Nitrosodimethylamine	0.400	0.357	10.8	81	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	86	0.00
5 S	Phenol-d6	0.400	0.369	7.8	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.364	9.0	87	0.00
9	Naphthalene	0.400	0.365	8.8	84	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	88	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	86	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.320	20.0	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	86	0.00
16	Acenaphthylene	0.400	0.342	14.5	85	0.00
17	Acenaphthene	0.400	0.359	10.3	86	0.00
18	Fluorene	0.400	0.359	10.3	86	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.000	100.0#	0	-15.63#
21	4-Bromophenyl-phenylether	0.400	0.361	9.8	89	0.00
22	Hexachlorobenzene	0.400	0.370	7.5	89	0.00
23	Atrazine	0.400	0.337	15.8	93	0.00
24	Pentachlorophenol	0.400	0.398	0.5	81	0.00
25	Phenanthrene	0.400	0.355	11.3	86	0.00
26	Anthracene	0.400	0.339	15.3	88	0.00
27 SURR	Fluoranthene-d10	0.400	0.352	12.0	90	0.00
28	Fluoranthene	0.400	0.350	12.5	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	95	0.00
30	Pyrene	0.400	0.368	8.0	92	0.00
31 S	Terphenyl-d14	0.400	0.358	10.5	87	0.00
32	Benzo(a)anthracene	0.400	0.347	13.3	95	0.00
33	Chrysene	0.400	0.368	8.0	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	112	0.00
37	Benzo(b)fluoranthene	0.400	0.340	15.0	102	0.00
38	Benzo(k)fluoranthene	0.400	0.338	15.5	102	0.00
39 C	Benzo(a)pyrene	0.400	0.356	11.0	113	0.00
40	Dibenzo(a,h)anthracene	0.400	0.359	10.3	113	0.00
41	Benzo(g,h,i)perylene	0.400	0.334	16.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN082922