

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082923\
 Data File : BN027246.D
 Acq On : 29 Aug 2023 17:59
 Operator : MA/JU
 Sample : 04126-13
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW203D-20230820

Quant Time: Aug 30 00:40:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	4075	0.400	ng	0.00
7) Naphthalene-d8	10.895	136	10795	0.400	ng	0.00
13) Acenaphthene-d10	14.694	164	7469	0.400	ng	0.00
19) Phenanthrene-d10	17.443	188	17157	0.400	ng	0.00
29) Chrysene-d12	21.623	240	13544	0.400	ng	0.00
35) Perylene-d12	24.153	264	12162	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.589	112	2226	0.210	ng	0.00
5) Phenol-d6	7.207	99	1515	0.117	ng	0.00
8) Nitrobenzene-d5	9.262	82	3942	0.501	ng	0.00
11) 2-Methylnaphthalene-d10	12.472	152	7973	0.503	ng	0.00
14) 2,4,6-Tribromophenol	16.189	330	1329	0.481	ng	0.00
15) 2-Fluorobiphenyl	13.325	172	12145	0.428	ng	0.00
27) Fluoranthene-d10	19.467	212	23744	0.557	ng	0.00
31) Terphenyl-d14	20.052	244	17068	0.627	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.465	88	1624	0.326	ng	# 88
34) Bis(2-ethylhexyl)phtha...	21.479	149	2874	0.128	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082923\
Data File : BN027246.D
Acq On : 29 Aug 2023 17:59
Operator : MA/JU
Sample : 04126-13
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW203D-20230820

Quant Time: Aug 30 00:40:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 29 05:37:30 2023
Response via : Initial Calibration

