



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082923\
 Data File : BN027231.D
 Acq On : 29 Aug 2023 08:45
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 29 05:37:30 2023






AutoFind: Scans 2536, 2537, 2538; Background Corrected with Scan 2529

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	149697	PASS
68	69	0.00	2	1.7	3059	PASS
69	198	0.00	100	43.1	175552	PASS
70	69	0.00	2	0.5	954	PASS
127	198	10	80	50.6	205888	PASS
197	198	0.00	2	0.3	1252	PASS
198	198	100	100	100.0	406912	PASS
199	198	5	9	6.9	27896	PASS
275	198	10	60	24.2	98312	PASS
365	198	1	100	2.8	11395	PASS
441	198	0.01	100	8.3	33667	PASS
442	442	50	100	100.0	224488	PASS
443	442	15	24	19.1	42955	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082923\
Data File : BN027231.D
Acq On : 29 Aug 2023 08:45
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 30 02:32:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 30 02:32:19 2023
Response via : Initial Calibration



TIC: BN027231.D\data.ms

(70) Pentachlorophenol (C)

17.069min (0.000) 21638.98 ng




response 446959

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.38
264.00	61.60	62.46
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082923\
Data File : BN027231.D
Acq On : 29 Aug 2023 08:45
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 30 02:32:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 30 02:32:19 2023
Response via : Initial Calibration



TIC: BN027231.D\data.ms

(77) Benzidine

19.698min (0.000) 0.00 ng

response 2925478

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.36
183.00	13.20	11.19
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/29/2023	BNA_N	<u>BN027231.D</u>
Compound Name	Response	Retention Time
DDT	1413227	20.892
DDD	16035	20.451
DDE	976	19.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
17011	1430238	1.19

Instrument :
BNA_N
ClientSampleId :
DFTPP