

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082924\
 Data File : BN033704.D
 Acq On : 30 Aug 2024 13:19
 Operator : MA/JU
 Sample : PB163069BSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB163069BSD

Quant Time: Aug 30 13:48:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 23:33:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.523	152	4730	0.400	ng	0.00
7) Naphthalene-d8	10.282	136	12128	0.400	ng	#-0.01
13) Acenaphthene-d10	14.157	164	5537	0.400	ng	-0.01
19) Phenanthrene-d10	16.917	188	10543	0.400	ng	0.00
29) Chrysene-d12	21.122	240	7773	0.400	ng	# 0.00
35) Perylene-d12	23.285	264	8210	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.169	112	4785	0.336	ng	0.00
5) Phenol-d6	6.714	99	5441	0.324	ng	0.00
8) Nitrobenzene-d5	8.659	82	4035	0.400	ng	0.00
11) 2-Methylnaphthalene-d10	11.881	152	7792	0.456	ng	0.00
14) 2,4,6-Tribromophenol	15.663	330	766	0.288	ng	0.00
15) 2-Fluorobiphenyl	12.778	172	9505	0.407	ng	0.00
27) Fluoranthene-d10	18.956	212	10005	0.385	ng	0.00
31) Terphenyl-d14	19.570	244	6933	0.422	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.168	88	1914	0.348	ng	# 42
3) n-Nitrosodimethylamine	3.457	42	2659	0.412	ng	# 97
6) bis(2-Chloroethyl)ether	6.960	93	5315	0.414	ng	98
9) Naphthalene	10.336	128	12856	0.396	ng	99
10) Hexachlorobutadiene	10.635	225	2665	0.396	ng	# 99
12) 2-Methylnaphthalene	11.956	142	8072	0.404	ng	99
16) Acenaphthylene	13.879	152	9797	0.409	ng	99
17) Acenaphthene	14.221	154	6768	0.406	ng	99
18) Fluorene	15.215	166	8096	0.392	ng	99
20) 4,6-Dinitro-2-methylph...	15.301	198	418	0.261	ng	96
21) 4-Bromophenyl-phenylether	16.123	248	2473	0.406	ng	93
22) Hexachlorobenzene	16.222	284	2898	0.416	ng	99
23) Atrazine	16.396	200	1903	0.392	ng	96
24) Pentachlorophenol	16.569	266	1545	0.491	ng	98
25) Phenanthrene	16.954	178	12055	0.411	ng	99
26) Anthracene	17.041	178	10590	0.417	ng	100
28) Fluoranthene	18.984	202	12723	0.385	ng	99
30) Pyrene	19.347	202	13119	0.417	ng	99
32) Benzo(a)anthracene	21.113	228	11786	0.430	ng	98
33) Chrysene	21.157	228	12063	0.430	ng	100
34) Bis(2-ethylhexyl)phtha...	21.068	149	10961	0.705	ng	# 69
36) Indeno(1,2,3-cd)pyrene	25.420	276	16185	0.459	ng	99
37) Benzo(b)fluoranthene	22.642	252	12962	0.430	ng	100
38) Benzo(k)fluoranthene	22.686	252	12501	0.422	ng	99
39) Benzo(a)pyrene	23.189	252	11323	0.449	ng	99
40) Dibenzo(a,h)anthracene	25.440	278	12744	0.454	ng	99
41) Benzo(g,h,i)perylene	26.075	276	13132	0.431	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082924\
Data File : BN033704.D
Acq On : 30 Aug 2024 13:19
Operator : MA/JU
Sample : PB163069BSD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Quant Time: Aug 30 13:48:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 23:33:13 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.523 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Instrument :
BNA_N
ClientSampleId :
PB163069BSD







Tgt Ion:152 Resp: 4730
Ion Ratio Lower Upper
152 100
150 157.7 123.4 185.0
115 65.3 51.3 76.9



#2
1,4-Dioxane
Concen: 0.348 ng
RT: 3.168 min Scan# 26
Delta R.T. -0.007 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion: 88 Resp: 1914
Ion Ratio Lower Upper
88 100
43 120.5 26.8 40.2#
58 95.3 63.4 95.0#









#3
 n-Nitrosodimethylamine
 Concen: 0.412 ng
 RT: 3.457 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19




Instrument :
 BNA_N
 ClientSampleId :
 PB163069BSD


Tgt Ion: 42 Resp: 2659
 Ion Ratio Lower Upper
 42 100
 74 112.1 91.8 137.6
 44 5.4 6.2 9.4#



#4
 2-Fluorophenol
 Concen: 0.336 ng
 RT: 5.169 min Scan# 303
 Delta R.T. 0.001 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19

Tgt Ion: 112 Resp: 4785
 Ion Ratio Lower Upper
 112 100
 64 60.8 48.6 72.8
 63 33.4 26.6 39.8









#5
Phenol-d6
Concen: 0.324 ng
RT: 6.714 min Scan# 518
Delta R.T. -0.007 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion: 99 Resp: 5441
Ion Ratio Lower Upper
99 100
42 22.8 17.8 26.8
71 35.5 28.0 42.0



#6
bis(2-Chloroethyl)ether
Concen: 0.414 ng
RT: 6.960 min Scan# 551
Delta R.T. -0.007 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion: 93 Resp: 5315
Ion Ratio Lower Upper
93 100
63 82.2 63.9 95.9
95 33.1 26.4 39.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.282 min Scan# 91
Delta R.T. -0.010 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:	136	Resp:	12128
Ion Ratio	Lower	Upper	
136	100		
137	11.5	9.1	13.7
54	9.6	5.8	8.6#
68	6.4	4.2	6.2#



#8
Nitrobenzene-d5
Concen: 0.400 ng
RT: 8.659 min Scan# 771
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:	82	Resp:	4035
Ion Ratio	Lower	Upper	
82	100		
128	41.5	29.3	43.9
54	56.1	49.4	74.0





#9

Naphthalene

Concen: 0.396 ng

RT: 10.336 min Scan# 91

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:128 Resp: 12856

Ion Ratio Lower Upper

128 100

129 11.5 9.4 14.0

127 13.4 11.0 16.4



#10

Hexachlorobutadiene

Concen: 0.396 ng

RT: 10.635 min Scan# 956

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:225 Resp: 2665


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.8 51.4 77.0









#11
 2-Methylnaphthalene-d10
 Concen: 0.456 ng
 RT: 11.881 min Scan# 1136
 Delta R.T. -0.007 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19

Instrument :
 BNA_N
 ClientSampleId :
 PB163069BSD

Tgt Ion:152 Resp: 7792




Ion	Ratio	Lower	Upper
152	100		
151	18.4	16.7	25.1




#12
 2-Methylnaphthalene
 Concen: 0.404 ng
 RT: 11.956 min Scan# 1156
 Delta R.T. -0.003 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19

Tgt Ion:142 Resp: 8072

Ion	Ratio	Lower	Upper
142	100		
141	88.4	71.8	107.8
115	38.9	31.4	47.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.157 min Scan# 1473
Delta R.T. -0.010 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:164 Resp: 5537
Ion Ratio Lower Upper
164 100
162 105.5 83.4 125.2
160 52.1 40.2 60.2



#14
2,4,6-Tribromophenol
Concen: 0.288 ng
RT: 15.663 min Scan# 1612
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:330 Resp: 766
Ion Ratio Lower Upper
330 100
332 96.6 78.6 118.0
141 42.8 35.0 52.6









#15
2-Fluorobiphenyl
Concen: 0.407 ng
RT: 12.778 min Scan# 1444
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:172 Resp: 9505
Ion Ratio Lower Upper
172 100
171 35.2 29.2 43.8
170 24.1 19.7 29.5



#16
Acenaphthylene
Concen: 0.409 ng
RT: 13.879 min Scan# 1447
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:152 Resp: 9797
Ion Ratio Lower Upper
152 100
151 20.0 15.6 23.4
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 0.406 ng
RT: 14.221 min Scan# 1480
Delta R.T. -0.010 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:154 Resp: 6768
Ion Ratio Lower Upper
154 100
153 115.1 91.8 137.8
152 58.7 45.4 68.2



#18
Fluorene
Concen: 0.392 ng
RT: 15.215 min Scan# 1572
Delta R.T. -0.010 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:166 Resp: 8096
Ion Ratio Lower Upper
166 100
165 99.3 78.6 118.0
167 13.8 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.917 min Scan# 11

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:188 Resp: 10543

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 7.8 7.4 11.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.261 ng

RT: 15.301 min Scan# 1580

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:198 Resp: 418


Ion Ratio Lower Upper

198 100

51 96.8 82.9 124.3

105 67.3 53.4 80.0









#21
4-Bromophenyl-phenylether
Concen: 0.406 ng
RT: 16.123 min Scan# 1649
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:	248	Resp:	2473
Ion Ratio	Lower	Upper	
248	100		
250	102.9	79.0	118.4
141	40.5	38.5	57.7



#22
Hexachlorobenzene
Concen: 0.416 ng
RT: 16.222 min Scan# 1657
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:	284	Resp:	2898
Ion Ratio	Lower	Upper	
284	100		
142	42.1	34.0	51.0
249	33.4	26.2	39.2





#23

Atrazine

Concen: 0.392 ng

RT: 16.396 min Scan# 1

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:200 Resp: 1903

Ion Ratio Lower Upper

200 100

173 31.2 27.0 40.4

215 46.5 35.5 53.3



#24

Pentachlorophenol

Concen: 0.491 ng

RT: 16.569 min Scan# 1685

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:266 Resp: 1545


Ion Ratio Lower Upper

266 100

264 62.6 51.4 77.2

268 66.2 51.9 77.9









#25
Phenanthrene
Concen: 0.411 ng
RT: 16.954 min Scan# 1716
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19




Instrument :
BNA_N
ClientSampleId :
PB163069BSD


Tgt Ion:178 Resp: 12055
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.2
179 15.6 12.2 18.2



#26
Anthracene
Concen: 0.417 ng
RT: 17.041 min Scan# 1723
Delta R.T. -0.012 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:178 Resp: 10590
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.4 12.2 18.4









#27
 Fluoranthene-d10
 Concen: 0.385 ng
 RT: 18.956 min Scan# 1938
 Delta R.T. 0.001 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19




Instrument : BNA_N
 ClientSampleId : PB163069BSD


Tgt Ion:	212	Resp:	10005
Ion Ratio	Lower	Upper	
212	100		
106	14.3	11.8	17.6
104	8.6	6.9	10.3



#28
 Fluoranthene
 Concen: 0.385 ng
 RT: 18.984 min Scan# 1944
 Delta R.T. -0.004 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19

Tgt Ion:	202	Resp:	12723
Ion Ratio	Lower	Upper	
202	100		
101	11.3	9.2	13.8
203	16.9	13.8	20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.122 min Scan# 24

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:240 Resp: 7773

Ion Ratio Lower Upper

240 100

120 20.3 13.3 19.9#

236 29.4 23.4 35.2



#30

Pyrene

Concen: 0.417 ng

RT: 19.347 min Scan# 2022

Delta R.T. -0.004 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:202 Resp: 13119


Ion Ratio Lower Upper

202 100

200 20.9 16.8 25.2

203 17.7 14.5 21.7









#31
 Terphenyl-d14
 Concen: 0.422 ng
 RT: 19.570 min Scan# 2070
 Delta R.T. 0.001 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19




Instrument :
 BNA_N
 ClientSampleId :
 PB163069BSD


Tgt Ion:244 Resp: 6933
 Ion Ratio Lower Upper
 244 100
 212 10.5 8.2 12.4
 122 16.4 13.6 20.4



#32
 Benzo(a)anthracene
 Concen: 0.430 ng
 RT: 21.113 min Scan# 2391
 Delta R.T. 0.010 min
 Lab File: BN033704.D
 Acq: 30 Aug 2024 13:19

Tgt Ion:228 Resp: 11786
 Ion Ratio Lower Upper
 228 100
 226 27.1 22.6 34.0
 229 20.2 15.8 23.6





#33

Chrysene

Concen: 0.430 ng

RT: 21.157 min Scan# 2396

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:228 Resp: 12063

Ion Ratio Lower Upper

228 100

226 31.3 24.7 37.1

229 19.6 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.705 ng

RT: 21.068 min Scan# 2386

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:149 Resp: 10961


Ion Ratio Lower Upper

149 100

167 15.6 28.3 42.5#

279 2.6 4.2 6.2#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.285 min Scan# 21
Delta R.T. 0.001 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Instrument :
BNA_N
ClientSampleId :
PB163069BSD

Tgt Ion:264 Resp: 8210




Ion	Ratio	Lower	Upper
264	100		
260	25.9	20.7	31.1
265	96.4	69.9	104.9




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.459 ng
RT: 25.420 min Scan# 3565
Delta R.T. -0.005 min
Lab File: BN033704.D
Acq: 30 Aug 2024 13:19

Tgt Ion:276 Resp: 16185

Ion	Ratio	Lower	Upper
276	100		
138	28.4	23.3	34.9
277	24.6	19.8	29.6





#37

Benzo(b)fluoranthene

Concen: 0.430 ng

RT: 22.642 min Scan# 2615

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:252 Resp: 12962

Ion Ratio Lower Upper

252 100

253 26.8 21.6 32.4

125 19.1 15.2 22.8



#38

Benzo(k)fluoranthene

Concen: 0.422 ng

RT: 22.686 min Scan# 2630

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:252 Resp: 12501


Ion Ratio Lower Upper

252 100

253 27.6 21.4 32.2

125 19.2 15.8 23.6





#39

Benzo(a)pyrene

Concen: 0.449 ng

RT: 23.189 min Scan# 21

Delta R.T. 0.001 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD





Tgt Ion:252 Resp: 11323

Ion Ratio Lower Upper

252 100

253 29.9 23.6 35.4

125 22.8 18.7 28.1



#40

Dibenzo(a,h)anthracene

Concen: 0.454 ng

RT: 25.440 min Scan# 3572

Delta R.T. 0.004 min

Lab File: BN033704.D

Acq: 30 Aug 2024 13:19




Tgt Ion:278 Resp: 12744


Ion Ratio Lower Upper

278 100

139 21.3 17.2 25.8

279 28.1 22.2 33.4





#41

Benzo(g,h,i)perylene

Concen: 0.431 ng

RT: 26.075 min Scan# 31

Delta R.T. 0.001 min

Lab File: BN033704.D


Acq: 30 Aug 2024 13:19

Instrument :

BNA_N

ClientSampleId :

PB163069BSD



Tgt Ion:276 Resp: 13132

Ion Ratio Lower Upper

276 100

277 25.6 20.2 30.2

138 27.0 21.1 31.7

