

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082924\  
 Data File : BN033695.D  
 Acq On : 30 Aug 2024 07:53  
 Operator : MA/JU  
 Sample : P3769-09  
 Misc :  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 OW-03B-51.5-082724-FD

Quant Time: Aug 30 08:24:43 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN082924.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 29 23:33:13 2024  
 Response via : Initial Calibration

Manual Integrations  
 APPROVED

Reviewed By :Jagrut Upadhyay 08/30/2024  
 Supervised By :mohammad ahmed 09/02/2024

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.530  | 152  | 5295     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.282 | 136  | 13683    | 0.400 | ng    | #-0.01   |
| 13) Acenaphthene-d10          | 14.167 | 164  | 6286     | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 16.917 | 188  | 12252    | 0.400 | ng    | 0.00     |
| 29) Chrysene-d12              | 21.121 | 240  | 9050     | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.285 | 264  | 9255     | 0.400 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.169  | 112  | 1852     | 0.116 | ng    | 0.00     |
| 5) Phenol-d6                  | 6.721  | 99   | 1348     | 0.072 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 8.659  | 82   | 3611     | 0.317 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 11.884 | 152  | 6253     | 0.324 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.663 | 330  | 851      | 0.282 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 12.777 | 172  | 8931     | 0.337 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 18.956 | 212  | 11961    | 0.396 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.569 | 244  | 8634     | 0.451 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| Target Compounds              |        |      |          |       |       | Qvalue   |
| 34) Bis(2-ethylhexyl)phtha... | 21.068 | 149  | 2168m    | 0.120 | ng    |          |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082924\  
Data File : BN033695.D  
Acq On : 30 Aug 2024 07:53  
Operator : MA/JU  
Sample : P3769-09  
Misc :  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
OW-03B-51.5-082724-FD

Quant Time: Aug 30 08:24:43 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN082924.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 29 23:33:13 2024  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Reviewed By :Jagrut Upadhyay 08/30/2024  
Supervised By :mohammad ahmed 09/02/2024

