




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083024\
Data File : BN033706.D
Acq On : 30 Aug 2024 14:31
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 04:51:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:51:24 2024
Response via : Initial Calibration



TIC: BN033706.D\data.ms

(70) Pentachlorophenol (C)




16.569min (0.000) 19020.84 ng

response 362728

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.40
264.00	61.60	63.65
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083024\
Data File : BN033706.D
Acq On : 30 Aug 2024 14:31
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 04:51:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:51:24 2024
Response via : Initial Calibration



TIC: BN033706.D\data.ms

(77) Benzidine

19.186min (0.000) 0.00 ng

response 2233991

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.56
183.00	13.20	11.79
0.00	0.00	0.00