




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083024\
Data File : BN033719.D
Acq On : 30 Aug 2024 23:25
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 04:52:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:51:24 2024
Response via : Initial Calibration



TIC: BN033719.D\data.ms

(70) Pentachlorophenol (C)




16.569min (-0.000) 20619.65 ng

response 428979

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.31
264.00	61.60	56.36
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083024\
Data File : BN033719.D
Acq On : 30 Aug 2024 23:25
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 04:52:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:51:24 2024
Response via : Initial Calibration



TIC: BN033719.D\data.ms

(77) Benzidine

19.180min (-0.006) 0.00 ng

response 1975039

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.47
183.00	13.20	11.60
0.00	0.00	0.00