

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083122\
 Data File : BN021486.D
 Acq On : 31 Aug 2022 11:00
 Operator : CG/JU
 Sample : PB147332BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB147332BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/01/2022
 Supervised By :Jagrut Upadhyay 09/01/2022

Quant Time: Sep 01 03:16:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.076	152	5117	0.400	ng	0.00
7) Naphthalene-d8	10.909	136	16182	0.400	ng	# 0.00
13) Acenaphthene-d10	14.729	164	8869	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	18417	0.400	ng	0.00
29) Chrysene-d12	21.664	240	12231	0.400	ng	# 0.00
35) Perylene-d12	24.188	264	9615	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.599	112	3588	0.358	ng	0.00
5) Phenol-d6	7.224	99	4599	0.360	ng	0.00
8) Nitrobenzene-d5	9.243	82	3311m	0.303	ng	-0.01
11) 2-Methylnaphthalene-d10	12.494	152	12382	0.339	ng	0.00
14) 2,4,6-Tribromophenol	16.219	330	820	0.282	ng	0.00
15) 2-Fluorobiphenyl	13.360	172	12496	0.322	ng	0.00
27) Fluoranthene-d10	19.506	212	16492	0.348	ng	0.00
31) Terphenyl-d14	20.097	244	12375	0.450	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083122\
Data File : BN021486.D
Acq On : 31 Aug 2022 11:00
Operator : CG/JU
Sample : PB147332BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB147332BL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/01/2022
Supervised By : Jagrut Upadhyay 09/01/2022

Quant Time: Sep 01 03:16:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 29 16:30:43 2022
Response via : Initial Calibration

