

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083122\
 Data File : BN021517.D
 Acq On : 01 Sep 2022 08:36
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 01 09:02:41 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	-0.01
2	1,4-Dioxane	0.400	0.434	-8.5	115	0.00
3	n-Nitrosodimethylamine	0.400	0.442	-10.5	118	0.00
4 S	2-Fluorophenol	0.400	0.442	-10.5	118	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	93	-0.01
8 S	Nitrobenzene-d5	0.400	0.404	-1.0	107	-0.01
9	Naphthalene	0.400	0.411	-2.7	103	-0.01
10	Hexachlorobutadiene	0.400	0.415	-3.7	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.447	-11.7	119	-0.01
12	2-Methylnaphthalene	0.400	0.443	-10.7	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.383	4.3	120	0.00
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	104	-0.01
16	Acenaphthylene	0.400	0.389	2.8	108	-0.01
17	Acenaphthene	0.400	0.409	-2.2	110	-0.01
18	Fluorene	0.400	0.413	-3.2	111	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.000	100.0#	0	-15.63#
21	4-Bromophenyl-phenylether	0.400	0.408	-2.0	112	-0.01
22	Hexachlorobenzene	0.400	0.412	-3.0	110	-0.01
23	Atrazine	0.400	0.386	3.5	119	0.00
24	Pentachlorophenol	0.400	0.458	-14.5	118	0.00
25	Phenanthrene	0.400	0.408	-2.0	110	-0.01
26	Anthracene	0.400	0.397	0.8	114	-0.01
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	114	0.00
28	Fluoranthene	0.400	0.397	0.8	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	95	0.00
30	Pyrene	0.400	0.446	-11.5	111	0.00
31 S	Terphenyl-d14	0.400	0.460	-15.0	112	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	110	0.00
33	Chrysene	0.400	0.421	-5.2	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.429	-7.2	118	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.413	-3.2	119	0.00
37	Benzo(b)fluoranthene	0.400	0.402	-0.5	110	0.00
38	Benzo(k)fluoranthene	0.400	0.398	0.5	110	0.00
39 C	Benzo(a)pyrene	0.400	0.413	-3.2	120	0.00
40	Dibenzo(a,h)anthracene	0.400	0.410	-2.5	118	-0.01
41	Benzo(g,h,i)perylene	0.400	0.396	1.0	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4