

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090121\
 Data File : BN016130.D
 Acq On : 01 Sep 2021 17:49
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4263

Quant Time: Sep 01 18:19:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 01 12:00:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.867	152	3245	0.400	ng/ul	0.00
4) Naphthalene-d8	10.666	136	13601	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.506	164	7651	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.250	188	15085	0.400	ng/ul	0.00
17) Chrysene-d12	21.442	240	13879	0.400	ng/ul	# 0.00
23) Perylene-d12	23.836	264	13707	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.319	96	1382	0.366	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.256	152	7943	0.415	ng/ul	0.00
18) Fluoranthene-d10	19.285	212	17059	0.458	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.353	88	1507	0.365	ng/ul#	86
5) Naphthalene	10.716	128	14052	0.388	ng/ul	99
7) 2-Methylnaphthalene	12.333	142	9558	0.405	ng/ul	99
8) 1-Methylnaphthalene	12.547	142	9475	0.402	ng/ul	100
10) Acenaphthylene	14.224	152	12543	0.433	ng/ul	99
11) Acenaphthene	14.566	153	9615	0.397	ng/ul	99
12) Fluorene	15.552	166	10865	0.392	ng/ul	100
14) Pentachlorophenol	16.904	266	1229	0.503	ng/ul	99
15) Phenanthrene	17.293	178	17047	0.392	ng/ul	99
16) Anthracene	17.386	178	16422	0.437	ng/ul	98
19) Fluoranthene	19.313	202	20123	0.435	ng/ul	99
20) Pyrene	19.675	202	20773	0.444	ng/ul	97
21) Benzo(a)anthracene	21.427	228	19410	0.457	ng/ul#	94
22) Chrysene	21.477	228	18925	0.395	ng/ul#	94
24) Benzo(b)fluoranthene	23.108	252	18805	0.365	ng/ul#	77
25) Benzo(k)fluoranthene	23.152	252	18605	0.359	ng/ul#	81
26) Benzo(a)pyrene	23.728	252	17557	0.401	ng/ul#	80
27) Indeno(1,2,3-cd)pyrene	26.320	276	21814	0.386	ng/ul#	94
28) Dibenzo(a,h)anthracene	26.337	278	18246	0.392	ng/ul#	80
29) Benzo(g,h,i)perylene	27.075	276	17987	0.367	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090121\
Data File : BN016130.D
Acq On : 01 Sep 2021 17:49
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4263

Quant Time: Sep 01 18:19:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 01 12:00:46 2021
Response via : Initial Calibration

