

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
 Data File : BN021531.D
 Acq On : 01 Sep 2022 18:38
 Operator : CG/JU
 Sample : N4297-14DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE131D2-20220821DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/02/2022
 Supervised By : Jagrut Upadhyay 09/02/2022

Quant Time: Sep 02 00:59:40 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	5277	0.400	ng	0.00
7) Naphthalene-d8	10.909	136	16489	0.400	ng	# 0.01
13) Acenaphthene-d10	14.728	164	8963	0.400	ng	0.01
19) Phenanthrene-d10	17.480	188	18746	0.400	ng	0.01
29) Chrysene-d12	21.664	240	11815	0.400	ng	# 0.00
35) Perylene-d12	24.188	264	7703	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.599	112	274	0.027	ng	0.00
5) Phenol-d6	7.231	99	183	0.014	ng	0.01
8) Nitrobenzene-d5	9.254	82	650m	0.058	ng	0.01
11) 2-Methylnaphthalene-d10	12.493	152	2377	0.064	ng	0.00
14) 2,4,6-Tribromophenol	16.218	330	159	0.054	ng	0.00
15) 2-Fluorobiphenyl	13.360	172	2463	0.063	ng	0.01
27) Fluoranthene-d10	19.506	212	3754	0.078	ng	0.00
31) Terphenyl-d14	20.097	244	2476	0.093	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	10834	1.642	ng	# 89
34) Bis(2-ethylhexyl)phtha...	21.557	149	446	0.027	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
Data File : BN021531.D
Acq On : 01 Sep 2022 18:38
Operator : CG/JU
Sample : N4297-14DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE131D2-20220821DL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/02/2022
Supervised By : Jagrut Upadhyay 09/02/2022

Quant Time: Sep 02 00:59:40 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

