

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
 Data File : BN021533.D
 Acq On : 01 Sep 2022 19:52
 Operator : CG/JU
 Sample : N4360-04DL 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D2-20220824DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/02/2022
 Supervised By : Jagrut Upadhyay 09/02/2022

Quant Time: Sep 02 01:00:23 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	5247	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	16469	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	9009	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	18788	0.400	ng	# 0.00
29) Chrysene-d12	21.664	240	11449	0.400	ng	0.00
35) Perylene-d12	24.182	264	7629	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.599	112	228	0.022	ng	0.00
5) Phenol-d6	7.224	99	172	0.013	ng	0.00
8) Nitrobenzene-d5	9.254	82	538m	0.048	ng	0.01
11) 2-Methylnaphthalene-d10	12.493	152	1797	0.048	ng	0.00
14) 2,4,6-Tribromophenol	16.229	330	81	0.027	ng	0.01
15) 2-Fluorobiphenyl	13.360	172	2884	0.073	ng	0.01
27) Fluoranthene-d10	19.506	212	2613	0.054	ng	0.00
31) Terphenyl-d14	20.097	244	1932	0.075	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	9693	1.477	ng	# 89
34) Bis(2-ethylhexyl)phtha...	21.557	149	465	0.029	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
Data File : BN021533.D
Acq On : 01 Sep 2022 19:52
Operator : CG/JU
Sample : N4360-04DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D2-20220824DL

Quant Time: Sep 02 01:00:23 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/02/2022
Supervised By : Jagrut Upadhyay 09/02/2022

