




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
 Data File : BN021545.D
 Acq On : 02 Sep 2022 04:28
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 02 06:54:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 06:51:31 2022
 Response via : Initial Calibration



TIC: BN021545.D\data.ms

(70) Pentachlorophenol (C)




17.116min (-0.006) 20295.75 ng

response 301350

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.03
264.00	61.60	61.36
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
Data File : BN021545.D
Acq On : 02 Sep 2022 04:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 02 06:54:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 06:51:31 2022
Response via : Initial Calibration



TIC: BN021545.D\data.ms

(77) Benzidine

19.716min (-0.006) 0.00 ng

response 1798157

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.29
183.00	13.20	11.29
0.00	0.00	0.00