

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090122\
 Data File : BN021557.D
 Acq On : 02 Sep 2022 11:46
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 03 00:08:51 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.500	0.542	-8.4	98	0.00
3	n-Nitrosodimethylamine	0.451	0.476	-5.5	96	0.00
4 S	2-Fluorophenol	0.783	0.846	-8.0	98	0.00
5 S	Phenol-d6	0.998	1.038	-4.0	97	0.00
6	bis(2-Chloroethyl)ether	1.256	1.305	-3.9	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.270	0.269	0.4	96	0.00
9	Naphthalene	1.185	1.219	-2.9	95	0.00
10	Hexachlorobutadiene	0.205	0.211	-2.9	94	-0.01
11 SURR	2-Methylnaphthalene-d10	0.902	1.062	-17.7	115	0.00
12	2-Methylnaphthalene	0.150	0.156	-4.0	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.131	0.123	6.1	110	-0.01
15 S	2-Fluorobiphenyl	1.749	1.783	-1.9	98	0.00
16	Acenaphthylene	1.878	1.829	2.6	102	0.00
17	Acenaphthene	1.321	1.332	-0.8	101	0.00
18	Fluorene	1.630	1.698	-4.2	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4,6-Dinitro-2-methylphenol	0.000	0.000	0.0	0#	-15.63#
21	4-Bromophenyl-phenylether	0.250	0.255	-2.0	104	0.00
22	Hexachlorobenzene	0.306	0.318	-3.9	104	0.00
23	Atrazine	0.146	0.135	7.5	105	-0.01
24	Pentachlorophenol	0.065	0.054	16.9	102	0.00
25	Phenanthrene	1.381	1.410	-2.1	102	0.00
26	Anthracene	1.097	1.081	1.5	106	0.00
27 SURR	Fluoranthene-d10	1.028	0.994	3.3	102	0.00
28	Fluoranthene	1.435	1.377	4.0	101	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
30	Pyrene	2.132	2.363	-10.8	98	0.00
31 S	Terphenyl-d14	0.899	1.047	-16.5	101	0.00
32	Benzo(a)anthracene	1.393	1.414	-1.5	99	0.00
33	Chrysene	1.659	1.710	-3.1	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.561	0.611	-8.9	106	0.00
35 I	Perylene-d12	1.000	1.000	0.0	86	0.00
36	Indeno(1,2,3-cd)pyrene	1.806	1.830	-1.3	103	-0.01
37	Benzo(b)fluoranthene	1.972	2.012	-2.0	99	0.00
38	Benzo(k)fluoranthene	1.984	1.945	2.0	96	0.00
39 C	Benzo(a)pyrene	1.376	1.460	-6.1	109	0.00
40	Dibenzo(a,h)anthracene	1.454	1.441	0.9	101	0.00
41	Benzo(g,h,i)perylene	1.595	1.560	2.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4