

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
 Data File : BN011689.D
 Acq On : 03 Sep 2020 09:16
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 03 12:35:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 03 12:33:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.643	0.586	8.9	89	0.00
3	n-Nitrosodimethylamine	0.743	0.734	1.2	88	0.00
4 S	2-Fluorophenol	1.130	1.001	11.4	83	0.00
5 S	Phenol-d6	1.535	1.319	14.1	83	0.00
6	bis(2-Chloroethyl)ether	1.249	1.135	9.1	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.387	0.359	7.2	87	0.00
9	Naphthalene	1.144	1.037	9.4	84	0.00
10	Hexachlorobutadiene	0.251	0.244	2.8	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.644	10.7	85	0.00
12	2-Methylnaphthalene	0.814	0.727	10.7	85	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.194	0.163	16.0	80	0.00
15 S	2-Fluorobiphenyl	1.691	1.584	6.3	86	0.00
16	Acenaphthylene	1.948	1.712	12.1	83	0.00
17	Acenaphthene	1.345	1.224	9.0	86	0.00
18	Fluorene	1.724	1.543	10.5	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.060	3.2	96	0.00
21	4-Bromophenyl-phenylether	0.234	0.213	9.0	85	0.00
22	Hexachlorobenzene	0.289	0.272	5.9	87	0.00
23	Atrazine	0.206	0.170	17.5	77	0.00
24	Pentachlorophenol	0.115	0.102	11.3	82	0.00
25	Phenanthrene	1.292	1.167	9.7	85	0.00
26	Anthracene	1.148	1.002	12.7	82	0.00
27 SURR	Fluoranthene-d10	1.284	1.117	13.0	81	0.00
28	Fluoranthene	1.532	1.341	12.5	82	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
30	Pyrene	1.391	1.339	3.7	82	0.00
31 S	Terphenyl-d14	0.993	0.964	2.9	82	0.00
32	Benzo(a)anthracene	1.428	1.292	9.5	78	0.00
33	Chrysene	1.473	1.390	5.6	80	0.00
34	Bis(2-ethylhexyl)phthalate	0.581	0.542	6.7	82	0.00
35	Indeno(1,2,3-cd)pyrene	1.921	1.759	8.4	76	0.00
36 I	Perylene-d12	1.000	1.000	0.0	77	0.00
37	Benzo(b)fluoranthene	1.534	1.353	11.8	80	0.00
38	Benzo(k)fluoranthene	1.568	1.438	8.3	79	0.00
39 C	Benzo(a)pyrene	1.371	1.187	13.4	78	0.00
40	Dibenzo(a,h)anthracene	1.548	1.338	13.6	75	0.00
41	Benzo(g,h,i)perylene	1.516	1.278	15.7	75	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
Data File : BN011689.D
Acq On : 03 Sep 2020 09:16
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 03 12:35:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 03 12:33:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0