

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
 Data File : BN011689.D
 Acq On : 03 Sep 2020 09:16
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 03 12:35:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 03 12:33:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.365	8.8	89	0.00
3	n-Nitrosodimethylamine	0.400	0.395	1.3	88	0.00
4 S	2-Fluorophenol	0.400	0.354	11.5	83	0.00
5 S	Phenol-d6	0.400	0.344	14.0	83	0.00
6	bis(2-Chloroethyl)ether	0.400	0.363	9.3	87	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	86	0.00
8 S	Nitrobenzene-d5	0.400	0.371	7.3	87	0.00
9	Naphthalene	0.400	0.362	9.5	84	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.357	10.8	85	0.00
12	2-Methylnaphthalene	0.400	0.357	10.8	85	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.335	16.3	80	0.00
15 S	2-Fluorobiphenyl	0.400	0.375	6.3	86	0.00
16	Acenaphthylene	0.400	0.352	12.0	83	0.00
17	Acenaphthene	0.400	0.364	9.0	86	0.00
18	Fluorene	0.400	0.358	10.5	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	84	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.391	2.3	96	0.00
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	85	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	87	0.00
23	Atrazine	0.400	0.331	17.3	77	0.00
24	Pentachlorophenol	0.400	0.355	11.3	82	0.00
25	Phenanthrene	0.400	0.361	9.8	85	0.00
26	Anthracene	0.400	0.349	12.8	82	0.00
27 SURR	Fluoranthene-d10	0.400	0.348	13.0	81	0.00
28	Fluoranthene	0.400	0.350	12.5	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
30	Pyrene	0.400	0.385	3.8	82	0.00
31 S	Terphenyl-d14	0.400	0.388	3.0	82	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	78	0.00
33	Chrysene	0.400	0.377	5.8	80	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.373	6.8	82	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	76	0.00
36 I	Perylene-d12	0.400	0.400	0.0	77	0.00
37	Benzo(b)fluoranthene	0.400	0.353	11.8	80	0.00
38	Benzo(k)fluoranthene	0.400	0.367	8.3	79	0.00
39 C	Benzo(a)pyrene	0.400	0.346	13.5	78	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	75	0.00
41	Benzo(g,h,i)perylene	0.400	0.337	15.8	75	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
Data File : BN011689.D
Acq On : 03 Sep 2020 09:16
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 03 12:35:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 03 12:33:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0