

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
 Data File : BN011696.D
 Acq On : 03 Sep 2020 14:08
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 03 14:39:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 03 12:33:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.643	0.561	12.8	81	0.00
3	n-Nitrosodimethylamine	0.743	0.813	-9.4	93	0.00
4 S	2-Fluorophenol	1.130	1.027	9.1	82	0.00
5 S	Phenol-d6	1.535	1.311	14.6	78	0.00
6	bis(2-Chloroethyl)ether	1.249	1.143	8.5	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
8 S	Nitrobenzene-d5	0.387	0.366	5.4	85	0.00
9	Naphthalene	1.144	1.036	9.4	81	0.01
10	Hexachlorobutadiene	0.251	0.242	3.6	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.647	10.3	82	0.00
12	2-Methylnaphthalene	0.814	0.732	10.1	82	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.194	0.170	12.4	82	0.00
15 S	2-Fluorobiphenyl	1.691	1.568	7.3	84	0.00
16	Acenaphthylene	1.948	1.712	12.1	81	0.00
17	Acenaphthene	1.345	1.240	7.8	85	0.01
18	Fluorene	1.724	1.538	10.8	82	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.063	-1.6	99	0.00
21	4-Bromophenyl-phenylether	0.234	0.219	6.4	86	0.00
22	Hexachlorobenzene	0.289	0.260	10.0	82	0.00
23	Atrazine	0.206	0.176	14.6	79	0.00
24	Pentachlorophenol	0.115	0.097	15.7	76	0.00
25	Phenanthrene	1.292	1.147	11.2	82	0.00
26	Anthracene	1.148	0.989	13.9	79	0.00
27 SURR	Fluoranthene-d10	1.284	1.074	16.4	77	0.00
28	Fluoranthene	1.532	1.298	15.3	78	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
30	Pyrene	1.391	1.397	-0.4	77	0.00
31 S	Terphenyl-d14	0.993	1.001	-0.8	77	0.00
32	Benzo(a)anthracene	1.428	1.304	8.7	71	0.00
33	Chrysene	1.473	1.399	5.0	73	0.00
34	Bis(2-ethylhexyl)phthalate	0.581	0.581	0.0	80	0.00
35	Indeno(1,2,3-cd)pyrene	1.921	1.749	9.0	69	0.00
36 I	Perylene-d12	1.000	1.000	0.0	70	0.00
37	Benzo(b)fluoranthene	1.534	1.333	13.1	71	0.00
38	Benzo(k)fluoranthene	1.568	1.439	8.2	72	0.00
39 C	Benzo(a)pyrene	1.371	1.189	13.3	71	0.00
40	Dibenzo(a,h)anthracene	1.548	1.338	13.6	68	0.00
41	Benzo(g,h,i)perylene	1.516	1.274	16.0	67	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090320\
Data File : BN011696.D
Acq On : 03 Sep 2020 14:08
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 03 14:39:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 03 12:33:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0