


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090324\
Data File : BN033733.D
Acq On : 03 Sep 2024 10:16
Operator : MA/JU
Sample : DFTPP232
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP232

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 04 01:39:37 2024



AutoFind: Scans 2447, 2448, 2449; Background Corrected with Scan 2440

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	38.8	107093	PASS
68	69	0.00	2	1.7	1863	PASS
69	198	0.00	100	40.5	111850	PASS
70	69	0.00	2	0.4	431	PASS
127	198	10	80	48.6	134355	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	276267	PASS
199	198	5	9	6.7	18608	PASS
275	198	10	60	26.5	73277	PASS
365	198	1	100	3.6	9936	PASS
441	198	0.01	100	10.8	29931	PASS
442	198	50	100	66.4	183531	PASS
443	442	15	24	18.6	34059	PASS

m/z	Abundance
39.00	327.0
40.00	1200.0
41.00	709.0
42.00	318.0
43.00	743.0
44.00	6185.0
44.80	307.0
48.90	398.0
50.90	203.0
55.00	517.0
56.00	242.0
56.90	476.0
66.90	209.0
68.00	184.0
69.00	465.0
70.00	232.0
71.00	221.0
72.90	572.0
74.90	152.0
77.00	329.0
79.00	174.0
82.10	223.0
83.10	274.0
83.90	439.0
85.00	183.0
86.00	177.0
96.00	251.0
97.00	183.0
111.00	162.0
118.80	167.0
133.00	282.0
147.10	219.0
177.00	175.0
190.90	261.0
192.80	268.0
207.00	2753.0
208.00	443.0
209.00	221.0
281.00	539.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
36.10	187.0
36.90	545.0
38.00	1450.0
39.00	8590.0
40.00	1714.0
41.00	658.0
43.00	716.0
44.00	5576.0
44.90	463.0
46.90	187.0
49.10	1059.0
50.00	29808.0
51.00	114088.0
52.00	5877.0
53.00	364.0
55.00	1024.0
56.00	3115.0
57.00	7787.0
57.90	457.0
58.90	219.0
59.90	241.0
61.00	1652.0
62.00	1557.0
63.00	4493.0
63.90	789.0
65.00	2319.0
65.90	197.0
67.00	305.0
68.00	2345.0
69.00	118480.0
69.90	634.0
70.90	229.0
73.00	1460.0
74.00	11703.0
75.00	19040.0
76.00	6059.0
77.00	130912.0
78.00	9900.0
79.00	8953.0
80.00	6661.0
81.00	9854.0
81.90	2506.0
83.00	1898.0
83.90	472.0
85.00	1668.0
86.00	3303.0
87.00	1285.0
87.90	490.0
89.00	379.0
91.00	2169.0
92.00	2880.0
93.00	14970.0
94.00	1277.0
94.90	436.0
96.00	945.0
97.10	416.0
98.00	11991.0
99.00	10491.0
99.90	1001.0
101.00	5050.0
101.90	272.0
103.00	1827.0
103.90	3270.0
104.90	3141.0
106.00	1035.0
107.00	38800.0
108.00	6243.0
109.00	1164.0
110.00	72424.0
111.00	11200.0
112.00	1371.0
113.00	384.0
115.00	280.0
116.00	2476.0
117.00	34920.0
118.00	2597.0
118.90	425.0
119.90	542.0
121.00	262.0
122.00	2775.0
123.00	4631.0
124.00	1963.0
124.90	1731.0
127.00	133760.0
128.00	10478.0
129.00	52928.0
130.00	5284.0
130.90	864.0
131.90	422.0
132.80	442.0
133.90	1469.0
135.00	4535.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

136.00	1697.0
137.10	2496.0
137.80	573.0
139.00	379.0
140.00	712.0
141.00	6771.0
142.00	2002.0
143.00	1376.0
144.00	499.0
144.90	434.0
146.00	1210.0
147.00	4069.0
148.00	8106.0
149.00	1856.0
150.00	608.0
151.00	822.0
151.80	520.0
152.90	2308.0
153.90	1486.0
155.00	3653.0
156.00	5488.0
157.00	1097.0
157.90	1396.0
158.90	936.0
160.00	2281.0
161.00	2907.0
162.00	963.0
163.00	242.0
163.90	385.0
165.00	2538.0
166.00	2150.0
167.00	13205.0
168.00	8925.0
169.00	1506.0
169.90	563.0
170.90	722.0
171.90	1251.0
173.00	1658.0
174.00	2692.0
175.00	5837.0
176.00	1080.0
176.90	2576.0
178.00	798.0
179.00	9930.0
180.00	8200.0
181.00	3473.0
182.20	354.0
183.00	280.0
183.90	690.0
185.00	4857.0
186.00	36480.0
187.00	9890.0
188.10	1469.0
189.00	2540.0
190.00	305.0
191.00	1530.0
192.00	3148.0
193.00	3684.0
194.00	788.0
195.00	337.0
195.90	7437.0
198.00	258496.0
199.00	17552.0
200.00	1405.0
201.60	957.0
203.00	2034.0
204.00	9368.0
205.00	17144.0
206.00	66400.0
207.00	10693.0
208.00	2730.0
209.00	1136.0
210.10	1334.0
210.90	2502.0
211.90	533.0
212.90	255.0
214.90	792.0
215.90	1472.0
217.00	18920.0
218.00	2439.0
219.10	246.0
221.00	8673.0
221.90	2684.0
223.00	4286.0
224.00	37136.0
225.00	9371.0
226.00	1119.0
227.00	15279.0
228.00	2037.0
229.00	3330.0
230.00	428.0
231.00	1425.0
233.10	245.0
234.00	974.0
235.00	1320.0
236.00	623.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

236.90	1927.0
237.90	242.0
239.00	750.0
240.00	666.0
241.00	894.0
242.00	1823.0
243.10	1960.0
244.00	26856.0
245.00	3674.0
246.00	6381.0
247.00	1319.0
248.00	260.0
248.90	1220.0
251.00	369.0
252.00	222.0
252.90	609.0
254.10	1197.0
255.00	143808.0
256.00	21000.0
257.00	2123.0
258.00	8294.0
258.90	1345.0
259.80	312.0
260.80	255.0
264.00	207.0
265.00	3452.0
265.90	538.0
267.10	154.0
270.00	193.0
271.00	452.0
272.00	294.0
273.00	4519.0
274.00	10576.0
275.00	64352.0
276.00	8881.0
277.00	5255.0
277.90	1363.0
278.90	300.0
281.00	632.0
282.00	302.0
283.00	685.0
283.90	443.0
284.90	796.0
286.10	183.0
289.00	160.0
290.00	202.0
291.00	164.0
291.90	320.0
293.00	1304.0
294.00	466.0
296.00	17640.0
296.90	2772.0
300.90	242.0
302.00	418.0
303.00	2292.0
304.10	541.0
308.00	246.0
308.90	359.0
309.80	256.0
313.00	236.0
314.00	753.0
314.90	1798.0
315.90	1168.0
317.00	283.0
321.00	722.0
321.90	255.0
323.00	5201.0
324.00	1083.0
326.00	162.0
326.90	1109.0
328.00	659.0
332.00	320.0
332.90	572.0
334.00	3907.0
335.10	818.0
341.00	680.0
341.90	211.0
345.90	1203.0
346.80	314.0
352.00	1607.0
353.00	1060.0
354.00	1710.0
355.10	321.0
359.00	154.0
365.00	8166.0
366.00	1040.0
371.00	337.0
372.00	2245.0
373.00	839.0
382.90	807.0
383.90	281.0
390.00	375.0
391.00	207.0
391.90	236.0
402.10	1434.0
402.90	1513.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

404.00	561.0
405.00	153.0
420.90	1281.0
421.90	941.0
423.00	9731.0
424.10	2729.0
424.90	349.0
441.00	22896.0
442.10	131840.0
443.10	25448.0
444.00	2495.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
35.90	158.0
37.00	603.0
38.00	1419.0
39.00	9683.0
39.90	1569.0
41.00	648.0
42.10	198.0
43.00	601.0
44.00	5326.0
44.90	440.0
49.10	1152.0
50.00	32216.0
51.00	130488.0
52.00	6834.0
52.90	286.0
55.00	1128.0
56.00	4290.0
57.00	9168.0
57.90	549.0
59.00	243.0
60.10	162.0
60.90	1752.0
61.90	1792.0
63.00	5138.0
64.00	898.0
65.10	2501.0
65.90	214.0
66.90	221.0
68.00	2419.0
69.00	136448.0
70.00	713.0
72.00	154.0
73.00	1610.0
74.00	13585.0
75.00	22616.0
76.00	7134.0
77.00	152576.0
78.00	10646.0
79.00	10146.0
80.00	8114.0
81.00	11073.0
82.00	2542.0
83.00	3026.0
83.90	654.0
84.90	2096.0
86.00	3235.0
87.00	1420.0
87.90	556.0
88.90	229.0
91.00	2596.0
92.00	2894.0
93.00	17632.0
94.00	1404.0
94.90	273.0
95.90	963.0
97.10	585.0
98.00	14204.0
99.00	11941.0
100.00	999.0
101.00	6485.0
102.00	398.0
102.90	2000.0
104.00	4016.0
105.00	3486.0
106.00	1254.0
107.00	48112.0
108.00	7274.0
109.00	1203.0
110.00	84760.0
111.00	13464.0
111.90	1843.0
112.90	438.0
114.90	401.0
116.00	2523.0
117.00	40264.0
118.00	3232.0
118.90	546.0
120.00	583.0
121.00	304.0
122.00	3338.0
123.00	4458.0
124.00	2161.0
125.00	2095.0
127.00	165888.0
128.00	11795.0
129.00	64520.0
130.00	5990.0
131.00	1190.0
131.90	573.0
132.90	470.0
134.00	2097.0
135.00	5487.0

Instrument :

BNA_N

ClientSampleId :

DFTPP232

136.00	2128.0
137.00	3114.0
137.80	676.0
138.80	463.0
139.90	995.0
141.00	8811.0
142.00	2911.0
142.90	1805.0
144.00	372.0
145.00	510.0
146.00	1713.0
147.00	4554.0
148.00	9965.0
148.90	2249.0
150.00	861.0
150.90	994.0
152.00	426.0
153.00	2583.0
154.00	2273.0
155.00	4698.0
156.00	6879.0
157.10	1339.0
158.00	1658.0
159.00	1223.0
160.00	2515.0
161.00	4329.0
161.90	1231.0
162.90	424.0
164.00	380.0
164.90	3415.0
166.00	2443.0
167.00	16824.0
167.90	9858.0
169.00	1285.0
170.00	704.0
170.90	760.0
171.90	1527.0
173.00	1992.0
174.00	3753.0
175.00	6445.0
176.00	1862.0
176.90	2893.0
177.90	1017.0
179.00	13450.0
180.00	9474.0
181.00	3889.0
182.00	703.0
182.90	312.0
183.90	1046.0
185.00	5984.0
186.00	45608.0
187.00	14234.0
188.00	1715.0
188.90	3150.0
189.90	519.0
191.00	1668.0
192.00	4091.0
193.00	4893.0
194.00	1038.0
194.90	478.0
195.90	8272.0
198.00	334016.0
199.00	23336.0
200.00	1910.0
201.50	1522.0
203.00	2702.0
204.00	12070.0
205.00	22328.0
206.00	92464.0
207.00	14468.0
208.00	3693.0
209.00	1314.0
209.90	1242.0
211.00	3617.0
211.90	779.0
212.90	292.0
215.00	1154.0
216.00	1738.0
217.00	23968.0
218.00	2982.0
218.90	376.0
221.00	11565.0
221.80	3330.0
223.00	5047.0
224.00	49336.0
225.00	12159.0
226.00	1384.0
227.00	21224.0
228.00	2980.0
229.00	4722.0
229.90	684.0
231.00	1856.0
231.90	218.0
232.70	378.0
234.00	1401.0
235.00	1748.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

236.00	1176.0
237.00	2501.0
238.00	277.0
239.00	937.0
240.00	733.0
241.00	1130.0
242.00	2534.0
243.00	2737.0
244.00	37728.0
245.00	4771.0
246.00	7953.0
247.00	1491.0
248.00	434.0
249.00	1385.0
250.00	278.0
251.00	472.0
251.90	514.0
252.90	819.0
255.00	201216.0
256.00	29640.0
257.00	2379.0
258.00	12053.0
259.00	1795.0
260.00	344.0
261.00	368.0
263.90	371.0
265.00	5047.0
265.90	794.0
267.00	186.0
270.00	183.0
270.90	652.0
272.00	573.0
273.00	5978.0
274.00	15588.0
275.00	89616.0
276.00	12645.0
277.00	7783.0
278.00	1246.0
278.90	217.0
281.00	545.0
282.00	340.0
283.10	868.0
284.00	511.0
285.00	1294.0
286.00	247.0
288.70	200.0
289.10	182.0
289.90	264.0
290.80	189.0
292.00	381.0
293.00	1934.0
293.90	484.0
295.00	413.0
296.00	24888.0
297.00	3259.0
298.00	281.0
301.00	301.0
302.00	415.0
303.00	2796.0
304.00	742.0
304.80	215.0
307.90	263.0
309.00	221.0
310.00	324.0
312.90	325.0
314.00	1424.0
315.00	3182.0
316.10	1298.0
317.00	342.0
321.00	682.0
322.00	320.0
323.10	7291.0
324.00	1256.0
324.90	231.0
325.90	165.0
327.00	1819.0
327.90	953.0
331.90	564.0
333.00	837.0
334.00	5312.0
335.00	1387.0
336.00	151.0
341.00	995.0
341.90	482.0
346.00	1838.0
346.90	372.0
350.90	207.0
352.00	2417.0
352.90	1756.0
354.00	2755.0
355.00	749.0
358.90	173.0
364.90	10892.0
365.90	1870.0
370.00	214.0
371.00	659.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

372.00	4292.0
373.10	1173.0
383.00	1145.0
384.00	383.0
389.90	473.0
391.00	383.0
391.90	385.0
401.10	252.0
402.00	1971.0
403.10	2336.0
403.90	1058.0
404.90	210.0
421.00	2191.0
422.00	2524.0
423.00	12966.0
424.00	4151.0
425.00	661.0
441.00	32520.0
442.00	232576.0
443.00	43328.0
444.00	4132.0
445.00	322.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

m/z	Abundance
37.00	529.0
38.00	913.0
39.10	6095.0
40.00	1507.0
41.10	666.0
42.10	158.0
43.00	586.0
44.00	5654.0
44.90	264.0
49.00	772.0
50.00	19648.0
51.00	77312.0
52.00	3660.0
53.00	322.0
55.00	837.0
55.90	2688.0
57.00	5872.0
58.00	306.0
58.90	185.0
60.10	193.0
60.90	1024.0
62.00	1250.0
63.00	3133.0
63.90	450.0
65.00	1567.0
65.90	159.0
67.00	380.0
68.00	1378.0
69.00	82016.0
70.00	643.0
72.90	1200.0
74.00	8887.0
75.00	13755.0
76.00	4418.0
77.00	91328.0
78.00	6601.0
79.00	6670.0
80.00	4634.0
81.00	7138.0
82.00	1520.0
83.00	1578.0
84.00	412.0
84.90	1455.0
86.00	2225.0
86.90	833.0
87.90	373.0
89.00	250.0
91.00	1933.0
92.00	1842.0
93.00	11254.0
94.00	956.0
94.90	226.0
96.10	778.0
97.20	301.0
98.00	8339.0
99.00	7418.0
100.00	601.0
101.00	3921.0
101.90	272.0
102.90	1546.0
103.90	2698.0
104.90	2401.0
106.00	914.0
107.00	30256.0
108.00	4893.0
108.90	780.0
110.00	55184.0
111.00	8289.0
112.00	1036.0
113.00	396.0
114.20	155.0
114.90	249.0
116.00	1709.0
117.00	26688.0
118.00	1917.0
118.90	313.0
120.00	441.0
120.90	213.0
122.00	1986.0
123.00	2881.0
123.90	1572.0
125.00	1480.0
127.00	103416.0
128.00	7729.0
129.00	41624.0
130.00	3369.0
130.90	709.0
132.00	458.0
132.80	475.0
133.90	1251.0
135.00	3529.0
135.90	1319.0

Instrument :

BNA_N

ClientSampleId :

DFTPP232

137.00	2125.0
137.90	287.0
139.90	613.0
141.00	5855.0
142.00	1655.0
142.90	1315.0
143.90	380.0
144.90	409.0
145.90	950.0
147.00	3100.0
148.00	6285.0
149.00	1714.0
150.00	272.0
151.10	726.0
153.00	1940.0
154.00	1249.0
155.00	3398.0
156.00	4587.0
157.00	984.0
158.00	958.0
159.00	763.0
160.00	1890.0
161.00	2474.0
161.90	757.0
163.00	329.0
164.00	374.0
165.00	2524.0
165.90	1791.0
167.00	11061.0
168.00	6961.0
169.00	961.0
169.90	390.0
170.90	544.0
171.90	1084.0
173.00	1473.0
174.00	2647.0
175.00	4634.0
176.00	1012.0
176.90	2187.0
177.80	851.0
178.90	8583.0
180.00	5887.0
181.00	2918.0
181.80	370.0
182.00	368.0
182.80	267.0
183.90	944.0
185.00	4381.0
186.00	31664.0
187.00	9231.0
188.00	1187.0
189.00	2266.0
189.90	389.0
190.90	1285.0
192.00	3086.0
193.00	3404.0
193.90	720.0
194.90	470.0
196.00	6419.0
198.00	236288.0
199.00	14937.0
199.90	1405.0
201.40	925.0
202.10	357.0
202.90	1573.0
204.00	9038.0
205.00	16536.0
206.00	64920.0
207.00	10415.0
208.00	2695.0
208.90	1337.0
210.10	990.0
211.10	2350.0
211.90	382.0
212.90	161.0
215.00	718.0
216.00	1510.0
217.00	18240.0
218.00	2106.0
218.90	222.0
220.90	8381.0
222.90	4209.0
224.00	36864.0
225.00	9600.0
226.00	1179.0
227.00	16245.0
228.00	2281.0
229.00	2717.0
230.00	553.0
231.00	1126.0
232.00	266.0
232.90	401.0
234.00	1023.0
235.00	1003.0
235.90	737.0
237.00	1360.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

237.80	235.0
238.90	609.0
240.00	569.0
240.90	662.0
242.00	1958.0
243.00	1929.0
244.00	28072.0
245.10	3307.0
246.00	5841.0
246.90	1212.0
248.00	288.0
248.90	1078.0
250.00	202.0
250.90	292.0
251.90	317.0
253.00	613.0
254.10	1213.0
255.00	154432.0
256.00	21232.0
257.00	1752.0
258.00	9023.0
259.00	1446.0
259.90	218.0
260.90	210.0
264.00	194.0
265.00	3518.0
266.00	379.0
267.00	232.0
269.90	183.0
270.90	541.0
271.90	396.0
273.00	5034.0
274.00	11520.0
275.00	65864.0
276.00	8808.0
277.00	6231.0
278.00	1196.0
278.90	246.0
281.00	609.0
282.10	410.0
283.00	949.0
284.00	495.0
285.00	1366.0
285.90	267.0
289.00	293.0
289.90	224.0
291.10	166.0
291.90	327.0
293.00	1369.0
294.10	369.0
294.90	527.0
296.00	21528.0
296.90	3026.0
297.90	269.0
301.00	201.0
302.00	413.0
303.00	2684.0
304.00	702.0
308.00	309.0
308.90	256.0
310.00	291.0
312.80	237.0
313.90	954.0
314.90	2427.0
316.00	1361.0
316.90	316.0
320.90	653.0
322.00	326.0
323.00	7262.0
324.00	1194.0
326.00	151.0
326.90	1358.0
328.00	828.0
331.90	559.0
333.00	478.0
334.00	5141.0
335.00	1268.0
341.00	964.0
341.90	240.0
345.90	1347.0
346.90	418.0
352.00	1965.0
353.00	1491.0
354.00	2208.0
355.10	415.0
358.90	220.0
365.00	10751.0
366.00	1578.0
369.90	217.0
370.90	373.0
372.00	3999.0
373.00	924.0
383.00	974.0
383.90	285.0
389.90	733.0
391.10	370.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232

391.90	272.0
401.00	269.0
402.00	1372.0
402.90	2479.0
403.90	746.0
421.00	1814.0
422.00	1698.0
423.00	12820.0
423.90	3286.0
425.10	388.0
441.00	34376.0
442.00	186176.0
443.00	33400.0
444.00	3094.0
445.10	220.0

Instrument :
BNA_N
ClientSampleId :
DFTPP232