




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090420\
Data File : BN011723.D
Acq On : 05 Sep 2020 05:00
Operator : CG/JU
Sample : L3843-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Quant Time: Sep 05 06:02:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN081220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 05 05:40:34 2020
Response via : Initial Calibration



TIC: BN011723.D

(10) Phenanthrene-d10 (I)

17.118min (-0.004) 0.40ng/ul m

response 539

Ion	Exp%	Act%
188.00	100	100
94.00	8.20	29.91#
80.00	9.00	30.79#
0.00	0.00	0.00