

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016195.D
 Acq On : 04 Sep 2021 16:14
 Operator : CG/JU
 Sample : PB138841BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138841BL

Quant Time: Sep 06 02:12:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.873	152	13512	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	71362	0.400	ng	0.00
13) Acenaphthene-d10	14.497	164	41299	0.400	ng	0.00
19) Phenanthrene-d10	17.249	188	81993	0.400	ng	0.00
29) Chrysene-d12	21.429	240	85008	0.400	ng	0.00
35) Perylene-d12	23.823	264	75787	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.475	112	12728	0.334	ng	0.02
5) Phenol-d6	7.043	99	15310	0.314	ng	0.00
8) Nitrobenzene-d5	9.025	82	19765	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.252	152	40020	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.994	330	4573	0.248	ng	0.00
15) 2-Fluorobiphenyl	13.127	172	55901	0.359	ng	0.00
27) Fluoranthene-d10	19.276	212	87029	0.353	ng	0.00
31) Terphenyl-d14	19.872	244	80012	0.372	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.301	93	36	0.001	ng	# 72
9) Naphthalene	10.711	128	366	0.002	ng	# 44
10) Hexachlorobutadiene	11.002	225	8	0.000	ng	# 54
12) 2-Methylnaphthalene	12.323	142	76	0.001	ng	# 56
16) Acenaphthylene	14.218	152	69	0.000	ng	# 42
17) Acenaphthene	14.560	154	34	0.000	ng	# 73
18) Fluorene	15.550	166	52	0.000	ng	# 85
20) 4,6-Dinitro-2-methylph...	15.690	198	11	0.286	ng	# 1
21) 4-Bromophenyl-phenylether	16.433	248	7	0.000	ng	# 54
22) Hexachlorobenzene	16.555	284	12	0.000	ng	# 7
23) Atrazine	16.701	200	11	0.000	ng	# 1
24) Pentachlorophenol	16.908	266	78	0.370	ng	# 78
25) Phenanthrene	17.285	178	172	0.001	ng	# 80
26) Anthracene	17.383	178	45	0.000	ng	# 55
28) Fluoranthene	19.306	202	63	0.000	ng	# 65
30) Pyrene	19.673	202	90	0.000	ng	# 48
32) Benzo(a)anthracene	21.429	228	421	0.001	ng	# 9
33) Chrysene	21.429	228	298	0.001	ng	# 14
34) Bis(2-ethylhexyl)phtha...	21.344	149	2139	0.011	ng	# 97
36) Indeno(1,2,3-cd)pyrene	26.284	276	8	0.000	ng	# 1
37) Benzo(b)fluoranthene	23.094	252	312	0.001	ng	# 1
38) Benzo(k)fluoranthene	23.149	252	56	0.000	ng	# 1
39) Benzo(a)pyrene	23.721	252	245	0.001	ng	# 1
40) Dibenzo(a,h)anthracene	26.322	278	19	0.000	ng	# 1
41) Benzo(g,h,i)perylene	27.048	276	18	0.000	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016195.D
 Acq On : 04 Sep 2021 16:14
 Operator : CG/JU
 Sample : PB138841BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138841BL

Quant Time: Sep 06 02:12:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.873 min Scan# 520
Delta R.T. -0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:152 Resp: 13512
Ion Ratio Lower Upper
152 100
150 152.9 125.3 187.9
115 59.3 48.4 72.6



#4
2-Fluorophenol
Concen: 0.334 ng
RT: 5.475 min Scan# 260
Delta R.T. 0.018 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:112 Resp: 12728
Ion Ratio Lower Upper
112 100
64 59.3 46.8 70.2
63 28.8 22.7 34.1









#5
Phenol-d6
Concen: 0.314 ng
RT: 7.043 min Scan# 430
Delta R.T. 0.009 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion: 99 Resp: 15310
Ion Ratio Lower Upper
99 100
42 16.4 13.8 20.6
71 30.0 24.3 36.5



#6
bis(2-Chloroethyl)ether
Concen: 0.001 ng
RT: 7.301 min Scan# 458
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion: 93 Resp: 36
Ion Ratio Lower Upper
93 100
63 72.2 62.8 94.2
95 77.8 27.0 40.4#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.660 min Scan# 753
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	136	Resp:	71362
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.9	13.3
54	9.0	7.2	10.8
68	6.4	5.2	7.8



#8
Nitrobenzene-d5
Concen: 0.345 ng
RT: 9.025 min Scan# 624
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	82	Resp:	19765
Ion Ratio	Lower	Upper	
82	100		
128	47.5	39.3	58.9
54	51.5	40.2	60.4









#9
Naphthalene
Concen: 0.002 ng
RT: 10.711 min Scan# 757
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	128	Resp:	366
Ion Ratio	Lower	Upper	
128	100		
129	37.3	8.9	13.3#
127	30.7	10.4	15.6#



#10
Hexachlorobutadiene
Concen: 0.000 ng
RT: 11.002 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	225	Resp:	8
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	100.0	51.2	76.8#









#11
2-Methylnaphthalene-d10
Concen: 0.352 ng
RT: 12.252 min Scan# 987
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:152 Resp: 40020
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.001 ng
RT: 12.323 min Scan# 1003
Delta R.T. -0.004 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:142 Resp: 76
Ion Ratio Lower Upper
142 100
141 104.5 70.4 105.6
115 96.4 24.4 36.6#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.497 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:164 Resp: 41299
Ion Ratio Lower Upper
164 100
162 99.3 79.8 119.6
160 43.2 34.2 51.4



#14
2,4,6-Tribromophenol
Concen: 0.248 ng
RT: 15.994 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:330 Resp: 4573
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.4 25.4 38.2









#15
2-Fluorobiphenyl
Concen: 0.359 ng
RT: 13.127 min Scan# 1093
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	172	Resp:	55901
Ion	Ratio	Lower	Upper
172	100		
171	34.0	26.9	40.3
170	21.7	17.1	25.7



#16
Acenaphthylene
Concen: 0.000 ng
RT: 14.218 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	152	Resp:	69
Ion	Ratio	Lower	Upper
152	100		
151	52.2	15.3	22.9#
153	27.5	10.4	15.6#









#17
Acenaphthene
Concen: 0.000 ng
RT: 14.560 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:154 Resp: 34
Ion Ratio Lower Upper
154 100
153 120.6 89.1 133.7
152 97.1 41.2 61.8#



#18
Fluorene
Concen: 0.000 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:166 Resp: 52
Ion Ratio Lower Upper
166 100
165 111.5 77.0 115.6
167 19.2 10.9 16.3#









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	188	Resp:	81993
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.0	5.0	7.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.286 ng
RT: 15.690 min Scan# 1295
Delta R.T. 0.051 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	198	Resp:	11
Ion Ratio	Lower	Upper	
198	100		
182	101.9	28.8	43.2#
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.000 ng
RT: 16.433 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:248 Resp: 7
Ion Ratio Lower Upper
248 100
250 105.7 74.6 112.0
141 137.7 51.6 77.4#



#22
Hexachlorobenzene
Concen: 0.000 ng
RT: 16.555 min Scan# 1365
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:284 Resp: 12
Ion Ratio Lower Upper
284 100
142 25.0 22.7 34.1
249 125.0 24.3 36.5#









#23
Atrazine
Concen: 0.000 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:200 Resp: 11
Ion Ratio Lower Upper
200 100
173 101.9 20.6 31.0#
215 109.3 40.5 60.7#



#24
Pentachlorophenol
Concen: 0.370 ng
RT: 16.908 min Scan# 1394
Delta R.T. -0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:266 Resp: 78
Ion Ratio Lower Upper
266 100
264 43.6 50.4 75.6#
268 47.4 49.6 74.4#









#25
Phenanthrene
Concen: 0.001 ng
RT: 17.285 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:178 Resp: 172
Ion Ratio Lower Upper
178 100
176 20.9 14.5 21.7
179 0.0 12.3 18.5#



#26
Anthracene
Concen: 0.000 ng
RT: 17.383 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:178 Resp: 45
Ion Ratio Lower Upper
178 100
176 40.0 13.8 20.8#
179 0.0 12.4 18.6#









#27
Fluoranthene-d10
Concen: 0.353 ng
RT: 19.276 min Scan# 1739
Delta R.T. -0.005 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:212 Resp: 87029
Ion Ratio Lower Upper
212 100
106 12.2 9.7 14.5
104 6.8 5.4 8.0



#28
Fluoranthene
Concen: 0.000 ng
RT: 19.306 min Scan# 1745
Delta R.T. -0.005 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:202 Resp: 63
Ion Ratio Lower Upper
202 100
101 0.0 8.4 12.6#
203 0.0 13.8 20.8#









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.429 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:240	Resp:	85008
Ion Ratio	Lower	Upper
240	100	
120	11.0	10.0 15.0
236	25.8	20.9 31.3



#30
Pyrene
Concen: 0.000 ng
RT: 19.673 min Scan# 1819
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:202	Resp:	90
Ion Ratio	Lower	Upper
202	100	
200	35.6	15.8 23.8#
203	48.9	14.1 21.1#









#31
 Terphenyl-d14
 Concen: 0.372 ng
 RT: 19.872 min Scan# 1859
 Delta R.T. 0.000 min
 Lab File: BN016195.D
 Acq: 04 Sep 2021 16:14




Instrument :
 BNA_N
 ClientSampleId :
 PB138841BL


Tgt Ion:244	Resp:	80012
Ion Ratio	Lower	Upper
244	100	
212	6.9	5.5 8.3
122	10.8	9.0 13.4



#32
 Benzo(a)anthracene
 Concen: 0.001 ng
 RT: 21.429 min Scan# 2020
 Delta R.T. 0.011 min
 Lab File: BN016195.D
 Acq: 04 Sep 2021 16:14

Tgt Ion:228	Resp:	421
Ion Ratio	Lower	Upper
228	100	
226	55.0	20.5 30.7#
229	82.1	16.2 24.4#









#33
Chrysene
Concen: 0.001 ng
RT: 21.429 min Scan# 2020
Delta R.T. -0.042 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	228	Resp:	298
Ion Ratio	Lower	Upper	
228	100		
226	55.0	22.5	33.7#
229	82.1	16.0	24.0#



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.011 ng
RT: 21.344 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14


Tgt Ion:	149	Resp:	2139
Ion Ratio	Lower	Upper	
149	100		
167	31.8	24.2	36.4
279	6.9	4.0	6.0#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.823 min Scan# 2526
Delta R.T. -0.003 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Instrument :
BNA_N
ClientSampleId :
PB138841BL







Tgt Ion:264 Resp: 75787
Ion Ratio Lower Upper
264 100
260 23.8 19.5 29.3
265 24.7 20.2 30.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.000 ng
RT: 26.284 min Scan# 3245
Delta R.T. -0.021 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:276 Resp: 8
Ion Ratio Lower Upper
276 100
138 0.0 19.5 29.3#
277 112.5 20.2 30.4#









#37
Benzo(b)fluoranthene
Concen: 0.001 ng
RT: 23.094 min Scan# 2313
Delta R.T. -0.003 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:252 Resp: 312
Ion Ratio Lower Upper
252 100
253 151.9 18.6 27.8#
125 340.5 10.5 15.7#



#38
Benzo(k)fluoranthene
Concen: 0.000 ng
RT: 23.149 min Scan# 2329
Delta R.T. 0.007 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:252 Resp: 56
Ion Ratio Lower Upper
252 100
253 140.4 18.2 27.2#
125 212.1 9.1 13.7#





#39

Benzo(a)pyrene

Concen: 0.001 ng

RT: 23.721 min Scan# 2496

Delta R.T. 0.003 min

Lab File: BN016195.D


Acq: 04 Sep 2021 16:14

Instrument :

BNA_N

ClientSampleId :

PB138841BL






Tgt Ion:252 Resp: 245

Ion Ratio Lower Upper

252 100

253 159.5 19.0 28.4#

125 333.1 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.000 ng

RT: 26.322 min Scan# 3256

Delta R.T. 0.000 min

Lab File: BN016195.D

Acq: 04 Sep 2021 16:14




Tgt Ion:278 Resp: 19


Ion Ratio Lower Upper

278 100

139 141.8 14.0 21.0#

279 157.2 20.2 30.2#





#41
 Benzo(g,h,i)perylene
 Concen: 0.000 ng
 RT: 27.048 min Scan# 3468
 Delta R.T. -0.017 min
 Lab File: BN016195.D
 Acq: 04 Sep 2021 16:14

Instrument :
 BNA_N
 ClientSampleId :
 PB138841BL

Tgt Ion:	276	Resp:	18
Ion Ratio	Lower	Upper	
276	100		
277	110.3	20.3	30.5#
138	95.3	17.5	26.3#

