

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016213.D
 Acq On : 05 Sep 2021 04:21
 Operator : CG/JU
 Sample : M3620-07
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-VPB-RW7-GW-680-682

Manual Integrations
 APPROVED

mohammad
 9/7/2021 11:40:28 AM

Quant Time: Sep 06 06:21:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.873	152	13762	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	74275	0.400	ng	# 0.00
13) Acenaphthene-d10	14.497	164	43177	0.400	ng	0.00
19) Phenanthrene-d10	17.236	188	87485	0.400	ng	#-0.01
29) Chrysene-d12	21.429	240	90360	0.400	ng	# 0.00
35) Perylene-d12	23.813	264	91111	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	3190	0.082	ng	0.00
5) Phenol-d6	7.034	99	2599	0.052	ng	0.00
8) Nitrobenzene-d5	9.025	82	15678	0.263	ng	0.00
11) 2-Methylnaphthalene-d10	12.247	152	30580	0.259	ng	0.00
14) 2,4,6-Tribromophenol	15.981	330	4912	0.255	ng	-0.01
15) 2-Fluorobiphenyl	13.114	172	44337	0.272	ng	-0.01
27) Fluoranthene-d10	19.271	212	81052	0.309	ng	0.00
31) Terphenyl-d14	19.867	244	81617	0.357	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.373	88	12238m	2.780	ng	
24) Pentachlorophenol	16.896	266	276	0.377	ng	# 84
34) Bis(2-ethylhexyl)phtha...	21.334	149	41470	0.203	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
Data File : BN016213.D
Acq On : 05 Sep 2021 04:21
Operator : CG/JU
Sample : M3620-07
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-VPB-RW7-GW-680-682

Manual Integrations
APPROVED

mohammad
9/7/2021 11:40:28 AM

Quant Time: Sep 06 06:21:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 06 02:08:58 2021
Response via : Initial Calibration

