

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016219.D
 Acq On : 05 Sep 2021 07:58
 Operator : CG/JU
 Sample : M3561-11
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUPLICATE-082521

Quant Time: Sep 06 06:22:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.873	152	15713	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	83909	0.400	ng	# 0.00
13) Acenaphthene-d10	14.497	164	65552	0.400	ng	# 0.00
19) Phenanthrene-d10	17.249	188	104998	0.400	ng	# 0.00
29) Chrysene-d12	21.440	240	110874	0.400	ng	# 0.01
35) Perylene-d12	23.837	264	111614	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.475	112	5720	0.129	ng	0.02
5) Phenol-d6	7.052	99	4433	0.078	ng	0.02
8) Nitrobenzene-d5	9.025	82	23165	0.344	ng	0.00
11) 2-Methylnaphthalene-d10	12.252	152	45715	0.342	ng	0.00
14) 2,4,6-Tribromophenol	15.994	330	9558	0.327	ng	0.00
15) 2-Fluorobiphenyl	13.127	172	65564	0.265	ng	0.00
27) Fluoranthene-d10	19.286	212	109907	0.349	ng	0.00
31) Terphenyl-d14	19.877	244	102349	0.365	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.711	128	5355	0.024	ng	# 20
17) Acenaphthene	14.560	154	57485	0.309	ng	# 90
18) Fluorene	15.550	166	44310	0.189	ng	# 28
28) Fluoranthene	19.316	202	12740	0.037	ng	# 89
30) Pyrene	19.678	202	39894	0.110	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.344	149	28998	0.116	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
Data File : BN016219.D
Acq On : 05 Sep 2021 07:58
Operator : CG/JU
Sample : M3561-11
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUPLICATE-082521

Quant Time: Sep 06 06:22:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 06 02:08:58 2021
Response via : Initial Calibration

