

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN090421\  
 Data File : BN016220.D  
 Acq On : 05 Sep 2021 08:34  
 Operator : CG/JU  
 Sample : M3561-13  
 Misc :  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 VE2-1

Quant Time: Sep 06 06:22:44 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN090421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Sep 06 02:08:58 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.873	152	15803	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	84470	0.400	ng	0.00
13) Acenaphthene-d10	14.497	164	52883	0.400	ng	0.00
19) Phenanthrene-d10	17.248	188	104199	0.400	ng	# 0.00
29) Chrysene-d12	21.439	240	104934	0.400	ng	# 0.01
35) Perylene-d12	23.833	264	110104	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.475	112	5382	0.121	ng	0.02
5) Phenol-d6	7.061	99	4070	0.071	ng	0.03
8) Nitrobenzene-d5	9.025	82	20648	0.304	ng	0.00
11) 2-Methylnaphthalene-d10	12.251	152	42727	0.318	ng	0.00
14) 2,4,6-Tribromophenol	15.994	330	8646	0.366	ng	0.00
15) 2-Fluorobiphenyl	13.127	172	62099	0.311	ng	0.00
27) Fluoranthene-d10	19.281	212	98637	0.315	ng	0.00
31) Terphenyl-d14	19.877	244	97518	0.368	ng	0.00
Target Compounds						
9) Naphthalene	10.723	128	6548	0.029	ng	# 51
17) Acenaphthene	14.560	154	21736	0.145	ng	98
18) Fluorene	15.550	166	24134	0.128	ng	# 90
26) Anthracene	17.370	178	27232	0.098	ng	# 80
30) Pyrene	19.673	202	8674	0.025	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.344	149	37971	0.160	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN090421\  
 Data File : BN016220.D  
 Acq On : 05 Sep 2021 08:34  
 Operator : CG/JU  
 Sample : M3561-13  
 Misc :  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampled :  
 VE2-1

Quant Time: Sep 06 06:22:44 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN090421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Sep 06 02:08:58 2021  
 Response via : Initial Calibration

